



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

gfc

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 12-1-2006 Cal 1 0.098 AC Cal 2 0.098 AC

The instrument was placed into service on Not placed into service.


Terry S. Gallegos

Terry S. Gallegos

1-4-2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 4, day of January, 2007.



Kathy A. Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

10f8
T04

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE *jk*

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 12-1-2006 TIME 1317

INTOXILYZER SERIAL # 80-000323 LOCATION Lab

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

Terry Gallegos

2058
TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
12/01/2006

je

13:19:07

Standard Lot#: 5026021
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:20:00
Diagnostic Test	Pass	13:20:27
Air Blank	0.000	13:20:55
0.100 Cal Check	0.099	13:21:13
Air Blank	0.000	13:21:43
Subject Test	0.000	13:22:00
Air Blank	0.000	13:22:27
Five Minute Wait	ABT*	
Air Blank	0.000	13:23:07

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
12/01/2006

je

13:23:39

Standard Lot#: 5026021
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	13:24:21
Air Blank	0.000	13:24:49

*Improper Sample

Not a Successfully Completed Test Sequence

4058
T24

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
12/01/2006

je

13:25:19

Standard Lot#: 5026021
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:26:23
Diagnostic Test	Pass	13:26:51
Air Blank	0.000	13:27:18
0.100 Cal Check	RFI*	13:27:29
Air Blank	0.000	13:27:59

*RFI Detect

Not a Successfully Completed Test Sequence

5098
T04

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
12/01/2006 13:28:06

Standard Lot#: 5026021
Last Changed By: T. GALLEGOS #28847
QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB
Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB
Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:29:05
Diagnostic Test	Pass	13:29:33
Air Blank	0.000	13:30:00
0.100 Cal Check	0.099	13:30:19
Air Blank	0.000	13:30:48
Subject Test	0.130*	13:33:57
Air Blank	0.000	13:34:29
Air Blank	0.000	13:34:56
Subject Test	XXX**	13:35:40
Air Blank	0.000	13:36:14

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6078
T24

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
12/01/2006 13:36:48

Standard Lot#: 5026021
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	13:37:15
Cal Check	0.098	13:37:34
Air Blank	0.000	13:38:03


Operator's Signature

70f8
TJ4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

file

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 12-1-2006 TIME 1317
Intoxilyzer Serial # 80-000323 Location Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE *Terry Gallegos*

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.


8 of 8
TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
12/01/2006 13:16:40

Standard Lot#: 5026021
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	13:17:08
Cal Check	0.098	13:17:26
Air Blank	0.000	13:17:56



Operator's Signature



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

je

1061

TS4

Intoxilyzer Serial Number: 80-000323

Initial Service Date: Not placed into service

CMI, Inc. Certificate of Calibration Date: 6.17.2003

11.13.2003

4.27.2005