



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

MS
5-25-07

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5-8-2007 Cal 1 0.100 AC Cal 2 0.098 AC

Between the dates of 3-27-2007 and 5-8-2007 the instrument was operating properly and accurately.


The instrument was taken out of service for repair on _____.
The instrument was returned to service on _____.

Terry S. Gallegos
Terry S. Gallegos

5-24-2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 24, day of May, 2007.



Kathy A. Kinney
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1077
B4
5/25/07

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallegos 20047 AGENCY Tucson Police Department

DATE 5.8.2007 TIME 1653

INTOXILYZER SERIAL # 80-000323 LOCATION Lab

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed ✓✓
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed ✓
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

Terry Gallegos

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
05/08/2007

16:55:13

*2077
PK
5-2507*

Standard Lot#: 627101G
Last Changed By: J. TRAINING #99999

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:56:15
Diagnostic Test	Pass	16:56:43
Air Blank	0.000	16:57:10
0.100 Cal Check	0.099	16:57:29
Air Blank	0.000	16:57:58
Subject Test	0.000	16:58:21
Air Blank	0.000	16:58:49
Five Minute Wait	ABT*	
Air Blank	0.000	16:59:25

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
05/08/2007 16:59:36

3 of 7
TS4
MMS
5-25-07

Standard Lot#: 627101G
Last Changed By: J. TRAINING #99999

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:00:27
Diagnostic Test	Pass	17:00:55
Air Blank	0.000	17:01:22
0.100 Cal Check	RFI*	17:01:38
Air Blank	IPS**	17:01:52

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000323
05/08/2007 17:02:04

4087
T34
5-25-07

Standard Lot#: 627101G
Last Changed By: J. TRAINING #99999

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:02:52
Diagnostic Test	Pass	17:03:20
Air Blank	0.000	17:03:47
0.100 Cal Check	0.099	17:04:06
Air Blank	0.000	17:04:35
Subject Test	0.008*	17:07:44
Air Blank	0.000	17:08:16
Air Blank	0.000	17:08:43
Subject Test	XXX**	17:09:10
Air Blank	0.000	17:09:43

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-000323

05/08/2007

17:10:04

Standard Lot#: 627101G

Last Changed By: J. TRAINING #99999

QAS: T. GALLEGOS #28847

TUCSON PD CRIME LAB

*Sof 7
T54 5-2007*

Test	g/210L	Time
Air Blank	0.000	17:10:31
Cal Check	0.098	17:10:50
Air Blank	0.000	17:11:20


Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6077
BY 11/19/07
52507

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 5.8.2007 TIME 1653
Intoxilyzer Serial # 80-000323 Location Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C + 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-000323

05/08/2007

16:53:26

7077
NMS
TS4
52507

Standard Lot#: 627101G

Last Changed By: J. TRAINING #99999

QAS: T. GALLEGOS #28847

TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	16:53:53
Cal Check	0.100	16:54:12
Air Blank	0.000	16:54:42


Operator's Signature