



CERTIFICATION by CUSTODIAN of RECORDS  
Tucson Police Department Crime Laboratory

TS4  
6-11-07

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 6-7-07 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 5-8-07 and 6-7-07 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.  
The instrument was returned to service on N/A.


Clare Conley

Clare L. Conley

6-8-2007

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 8, day of June, 2007.



Kathy A Kinney

Notary Public

\*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Clare Conley 34517 AGENCY Tucson Police Department  
DATE 6-7-07 TIME 0635  
Intoxilyzer Serial # 80-000323 Location TPD Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

2 of 8 cu

784  
6.11.07

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-000323  
06/07/2007 06:35:51

-----  
Standard Lot#: 627101G  
Last Changed By: J. TRAINING #99999  
-----

QAS: C. CONLEY #34517  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	06:36:18
Cal Check	0.099	06:36:37
Air Blank	0.000	06:37:07

*Clare Conley*

\_\_\_\_\_  
Operator's Signature

3 of 8 04

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

TSK  
6-11-07

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clare Conley 34517 AGENCY Tucson Police Department

DATE 6-7-07 TIME 0635

INTOXILYZER SERIAL # 80-000323 LOCATION TPD Crime Laboratory

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ( ✓ ) 1. Clock time check.
- ( ✓ ) 2. Date check.

OPERATIONAL TESTS

- ( ✓ ) 1. Alcohol-free subject test result 0.000 AC.
- ( ✓ ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- ( ✓ ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- ( ✓ ) 4. Standard Calibration Check standard 0.100 AC  
Result: 0.098 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

---



---



---



---

SIGNATURE Clare Conley

4 of 8 Cu

By  
6-11-07

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-000323  
06/07/2007 06:38:09

---

Standard Lot#: 627101G  
Last Changed By: J. TRAINING #99999

---

QAS: C. CONLEY #34517  
TUCSON PD

---

Operator: C. CONLEY #34517  
TUCSON PD

---

Subject: TEST,TEST  
DOB: 11/11/1111  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	06:39:36
Diagnostic Test	Pass	06:40:04
Air Blank	0.000	06:40:31
0.100 Cal Check	0.099	06:40:50
Air Blank	0.000	06:41:19
Subject Test	0.000	06:41:37
Air Blank	0.000	06:42:05
Five Minute Wait	ABT*	
Air Blank	0.000	06:42:46

\*Sequence Aborted

### Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 8 cu

Ph  
6-11-07

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-000323  
06/07/2007 06:43:46

Standard Lot#: 627101G  
Last Changed By: J. TRAINING #99999

QAS: C. CONLEY #34517  
TUCSON PD

Operator: C. CONLEY #34517  
TUCSON PD

Subject: TEST,TEST  
DOB: 11/11/1111  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	06:45:04
Diagnostic Test	Pass	06:45:31
Air Blank	0.000	06:45:59
0.100 Cal Check	0.098	06:46:18
Air Blank	0.000	06:46:47
Subject Test	RFI*	06:46:58
Air Blank	0.000	06:47:26

\*RFI Detect

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

b of 8 Cu

Tdy  
6.11.07

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-000323  
06/07/2007 06:47:43

---

Standard Lot#: 627101G  
Last Changed By: J. TRAINING #99999

---

QAS: C. CONLEY #34517  
TUCSON PD

---

Operator: C. CONLEY #34517  
TUCSON PD

---

Subject: TEST,TEST  
DOB: 11/11/1111  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	06:48:37
Diagnostic Test	Pass	06:49:05
Air Blank	0.000	06:49:32
0.100 Cal Check	0.099	06:49:51
Air Blank	IPS*	06:49:58
Air Blank	0.000	06:50:26

\*Improper Sample

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 8 Cu

TJ4  
6-11-07

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-000323  
06/07/2007 06:50:32

Standard Lot#: 627101G  
Last Changed By: J. TRAINING #99999

QAS: C. CONLEY #34517  
TUCSON PD

Operator: C. CONLEY #34517  
TUCSON PD

Subject: TEST,TEST  
DOB: 11/11/1111  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	06:51:27
Diagnostic Test	Pass	06:51:55
Air Blank	0.000	06:52:22
0.100 Cal Check	0.099	06:52:41
Air Blank	0.000	06:53:10
Subject Test	0.000*	06:56:19
Air Blank	0.000	06:56:46
Air Blank	0.000	06:57:13
Subject Test	XXX**	06:57:34
Air Blank	0.000	06:58:08

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

### Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

8 of 8 cu

BY  
6-11-07

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-000323  
06/07/2007 06:58:38

-----  
Standard Lot#: 627101G  
Last Changed By: J. TRAINING #99999  
-----

QAS: C. CONLEY #34517  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	06:59:05
Cal Check	0.098	06:59:24
Air Blank	0.000	06:59:54

*Clare Conley*

\_\_\_\_\_  
Operator's Signature