



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TB4
8.25.2015

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 8/20/2015 Cal 1 0.096 AC Cal 2 0.096 AC

Between the dates of 7/23/2015 and 8/20/2015 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on .
The instrument was returned to service on .

Amy L. Autret
Amy L. Autret


8/24/2015

ADPS QAS Permit # 33718

SUBSCRIBED AND SWORN to before me this 24th, day of August, 2015.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1087
db
TJG
8-25-2015

QA SPECIALIST Amy Autret 53546 AGENCY Tucson Police Department

DATE 8/20/2015 TIME 1354

INTOXILYZER SERIAL # 80-000323 LOCATION Crime Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.


OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 096 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE



Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2017
Alb
Bh

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
08/20/2015 13:57:42

Standard Lot#: 25913100A2
Last Changed By: T. GALLEGOS #28847

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SOAP, SOAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:59:07
Diagnostic Test	Pass	13:59:43
Air Blank	0.000	14:00:11
0.100 Cal Check	0.096	14:00:31
Air Blank	0.000	14:01:01
Subject Test	0.000	14:01:26
Air Blank	0.000	14:01:53
Wait	ABT*	
Air Blank	0.000	14:02:28

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

3 of 7
ML
TB4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
08/20/2015 14:03:27

Standard Lot#: 25913100A2
Last Changed By: T. GALLEGOS #28847

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SOAP, SOAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:04:42
Diagnostic Test	Pass	14:05:18
Air Blank	0.000	14:05:46
0.100 Cal Check	RFI*	14:05:59
Air Blank	IPS**	14:06:10

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

4 of 7
AA
754

INTECHLYZER 8010
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
08/20/2015 14:07:20

Standard Lot#: 259:3100A2
Last Changed By: T. GALLEGUS #28847

OAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SOAP,SOAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:08:33
Diagnostic Test	Pass	14:09:09
Air Blank	0.000	14:09:37
0.100 Cal Check	0.096	14:09:58
Air Blank	0.000	14:10:27
Subject Test	0.135*	14:13:35
Air Blank	0.000	14:14:07
Air Blank	0.000	14:14:34
Subject Test	XXX**	14:15:10
Air Blank	0.000	14:15:44

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7
JL
TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
08/20/2015 14:16:33

Standard Lot#: 25913100A2
Last Changed By: T. GALLEGOS #28847

OAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	14:17:13
0.100 Cal Check	0.096	14:17:24
Air Blank	0.000	14:17:53

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
TS4
8.25.2015

QA Specialist Amy Autret 53546 AGENCY Tucson Police Department
DATE 8/20/2015 TIME 1354
Intoxilyzer Serial # 80-000323 Location Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 096 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE



DPS Form Exh G-2 (Rev. 05-01)

7067
JL
TS4

INTOXILYZER 8000
Instrument Initialization
09:41:26 08/20/2015

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
08/20/2015 13:54:29

Standard Lot#: 25913100A2
Last Changed By: T. GALLEGOS #28847

CAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:55:10
0.100 Cal Check	0.096	13:55:31
Air Blank	0.000	13:56:00

Preventative Maintenance
Performed -- Timer Reset