



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 1306 W. Miracle Mile, Tucson, AZ 85705

Old
 10/5/15

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000323 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 9.21.2015 Cal 1 0.095 AC Cal 2 0.095 AC

Between the dates of 8.20.2015 and 9.21.2015 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
 The instrument was returned to service on —.


Terry Gallegos

Terry S. Gallegos

9-29-2015

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 29th, day of September, 20 15.



Elma Escarcega

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
TS4
10/5/15

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 9.21.2015 TIME 1542

INTOXILYZER SERIAL # 80-000323 LOCATION Crime Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 095 AC

Instrument is operating properly and accurately. YES NO

COMMENTS > 31 days for calibration checks

SIGNATURE Terry Gallegos

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7
TS4
JLS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
09/21/2015 15:46:47

Standard Lot#: 25913100A2
Last Changed By: T. GALLEGOS #28847

OAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:49:36
Diagnostic Test	Pass	15:50:13
Air Blank	0.000	15:50:40
0.100 Cal Check	0.096	15:51:01
Air Blank	0.000	15:51:31
Subject Test	0.000	15:52:15
Air Blank	0.000	15:52:42
Wait	ABT*	
Air Blank	0.000	15:53:22

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

3 of 7
TS4
JL

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
09/21/2015 15:54:37

Standard Lot#: 25913100A2
Last Changed By: T. GALLEGOS #28847

OAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SOAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:55:59
Diagnostic Test	Pass	15:56:36
Air Blank	0.000	15:57:03
0.100 Cal Check	RFI*	15:57:23
Air Blank	IPS**	15:57:33

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

4 of 7
TS4
JLD

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
09/21/2015 16:00:58

Standard Lot#: 259131 00A2
Last Changed By: T. GALLEGOS #28847

DAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:02:53
Diagnostic Test	Pass	16:03:30
Air Blank	0.000	16:03:57
0.100 Cal Check	0.096	16:04:18
Air Blank	0.000	16:04:47
Subject Test	0.035*	16:07:55
Air Blank	0.000	16:08:26
Air Blank	0.000	16:08:53
Subject Test	XXX**	16:09:22
Air Blank	0.000	16:09:56

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7
TS4
JLB

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000323
Core Version: 8105.48
09/21/2015 16:10:27

Standard Lot#: 25913100A2
Last Changed By: T. GALLEGOS #28847

OAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	16:10:56
0.100 Cal Check	0.095	16:11:17
Air Blank	0.000	16:11:46

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

60f7
TSG
All
10/5/15

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 9-21-2015 TIME 1542
Intoxilyzer Serial # 80-000323 Location Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "~~PUSH BUTTON TO START~~" ^{TSG}
"Contact QAS 31 Day Pre-maint"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 095 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

DPS Form Exh G-2 (Rev. 05-01)

7 of 7
TS4
[Signature]

INTOXILYZER 8000
Instrument Initialization
14:33:36 09/21/2015

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-00323
Core Version: 8105.48
09/21/2015 15:42:16

Standard Lot#: 25913100A2
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	15:42:57
0.100 Cal Check	0.095	15:43:17
Air Blank	0.000	15:43:47

Preventative Maintenance
Performed -- Timer Reset
[Arrow pointing to 'Timer Reset']