CU
10-25-07

CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000331 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 9-27-2007 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 8-30-2007 and 9-27-2007 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Terry Gallegos

Terry S. Gallegos

10-25-2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 25, day of October, 2007.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

Ch 10-25-07

1 of 8

TSH

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 9-27-2007 TIME 1011

INTOXILYZER SERIAL # 80-000331 LOCATION ODW

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed ✓✓✓
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.100 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Terry Gallegos

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

lh

2 of 8

TJ4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000331
09/27/2007 10:13:23

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:14:09
Diagnostic Test	Pass	10:14:37
Air Blank	0.000	10:15:04
0.100 Cal Check	0.101	10:15:23
Air Blank	0.000	10:15:54
Subject Test	0.000	10:16:19
Air Blank	0.000	10:16:46
Five Minute Wait	ABT*	
Air Blank	0.000	10:17:21

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Cu
3 of 8
TSY

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000331
09/27/2007 10:17:44

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:18:29
Diagnostic Test	Pass	10:18:57
Air Blank	0.000	10:19:24
0.100 Cal Check	RFI*	10:19:39
Air Blank	IPS**	10:19:51

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Cu

4 of 8

P4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000331
09/27/2007 10:20:11

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:20:59
Diagnostic Test	Pass	10:21:27
Air Blank	0.000	10:21:54
0.100 Cal Check	0.101	10:22:13
Air Blank	0.000	10:22:43
Subject Test	RFI*	10:23:30
Air Blank	0.000	10:24:03

*RFI Detect

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

Ch

5 of 8
TOH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000331
09/27/2007 10:24:23

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:25:09
Diagnostic Test	Pass	10:25:36
Air Blank	0.000	10:26:04
0.100 Cal Check	0.101	10:26:22
Air Blank	0.000	10:26:53
Subject Test	0.493*	10:30:01
Air Blank	0.000	10:30:34
Air Blank	0.000	10:31:01
Subject Test	XXX**	10:31:25
Air Blank	0.000	10:31:59

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000331
09/27/2007 10:32:21

Ch
6 of 8
TS4

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	10:32:48
Cal Check	0.100	10:33:07
Air Blank	0.000	10:33:38

Jerry Galley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

Ch 10.25.07
7 of 8
PG

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 9-27-2007 TIME 1011

Intoxilyzer Serial # 80-000331 Location ODW

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gally

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000331
09/27/2007 10:11:28

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	10:11:55
Cal Check	0.101	10:12:14
Air Blank	0.000	10:12:44

Operator's Signature

Ch
8 of 8
B4

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.