CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000331 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

BY 4.11.2008

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4/10/08 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 3/13/08 and 4/10/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Sarah Cione

Sarah L. Cione

4/11/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 11, day of April, 2008.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

BY 4-11-2008

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 4/10/08 TIME 9:51

INTOXILYZER SERIAL # 80-000331 LOCATION Team 2 - DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.099 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Sarah Cione

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7sc

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000331
04/10/2008 09:53:34

TSH

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE
TUCSON PD #50616

Operator: S. CIONE
TUCSON PD #50616

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:54:30
Diagnostic Test	Pass	09:54:58
Air Blank	0.000	09:55:25
0.100 Cal Check	0.101	09:55:44
Air Blank	0.000	09:56:13
Subject Test	0.000	09:56:30
Air Blank	0.000	09:56:58
Five Minute Wait	ABT*	
Air Blank	0.000	09:57:34

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7 SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000331
04/10/2008 09:57:47

134

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE
TUCSON PD #50616

Operator: S. CIONE
TUCSON PD #50616

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	09:58:51
Air Blank	IPS**	09:59:03

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7 SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000331
04/10/2008 09:59:47

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE
TUCSON PD #50616

Operator: S. CIONE
TUCSON PD #50616

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:00:35
Diagnostic Test	Pass	10:01:02
Air Blank	0.000	10:01:29
0.100 Cal Check	0.100	10:01:48
Air Blank	0.000	10:02:18
Subject Test	0.095*	10:05:26
Air Blank	0.000	10:05:58
Air Blank	0.000	10:06:25
Subject Test	XXX**	10:06:52
Air Blank	0.000	10:07:25

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000331
04/10/2008
10:10:02

5 of 7 SC

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

B4

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:10:30
Cal Check	0.099	10:10:48
Air Blank	0.000	10:11:18

Sarah Cione
Operator's Signature

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

6 of 7 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

BH 4.11.2008

QA Specialist Sarah Cione AGENCY Tucson Police Department

DATE 4/10/08 TIME 9:51

Intoxilyzer Serial # 80--000331 Location Team 2 - DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000331
04/10/2008 09:51:45

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:52:12
Cal Check	0.100	09:52:31
Air Blank	0.000	09:53:01

Sarah Cione
Operator's Signature

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.