CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSH 4-12-2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000331 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4/1/10 Cal 1 0.102 AC Cal 2 0.101 AC

Between the dates of 3/4/10 and 4/1/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on –.
The instrument was returned to service on –.

Sarah Cione

Sarah L. Cione

4/12/10

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 12th, day of April, 2010.

AM Hernández

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

1 of 7 sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSY
4-12-2010

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 4/1/10 TIME 1025

INTOXILYZER SERIAL # 80-000331 LOCATION vehicle #432

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.101 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Sarah Cione

DPS form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7 sc

184

INTOXILYZER 5000
Location: TUCSON POLICE DEPT
Serial Number: 80-000331
04/01/2010 10:27:13

Standard Lot#: 054582
Last Changed By: C. CONLEY #38517

OAS: S. CIDNE #50616
TUCSON PD

Operator: S. CIDNE #50616
TUCSON PD

Subject: SOAP,SOAP
DOB: 11/11/1111
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:27:57
Diagnostic Test	Pass	10:28:25
Air Blank	0.000	10:28:53
0.100 Cal Check	0.099	10:29:11
Air Blank	0.000	10:29:41
Subject Test	0.000	10:29:55
Air Blank	0.000	10:30:22
Five Minute Wait	ABT*	
Air Blank	0.000	10:30:56

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7 sc

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 81-000331
04/01/2010 10:31:03

Standard Lot#: 654692
Last Changed By: C. CONLEY #34517

DMS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SDAP,SDAP
DOB: 11/11/1111
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:31:41
Air Blank	IPS**	10:31:49

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7 sc

INTOXILYZER 8100
Location: TUCSON POLICE DEPT
Serial Number: 80-100331
04/01/2010 10:31:56

Standard Lot#: 054582
Last Changed By: C. CONLEY #34517

DRE: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP,SOAP
DOB: 11/11/1111
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:32:50
Diagnostic Test	Pass	10:33:19
Air Blank	0.000	10:33:46
0.100 Cal Check	0.099	10:34:05
Air Blank	0.000	10:34:35
Subject Test	0.156*	10:37:43
Air Blank	0.000	10:38:15
Air Blank	0.000	10:38:42
Subject Test	XXX**	10:38:58
Air Blank	0.000	10:39:31

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7 sc

pg4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 88-000331
04/01/2010 10:40:19

Standard Lot#: 654682
Last Changed By: C. CONLEY #34517

OPS: S. CIONE #50516
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:40:46
Cal Check	0.101	10:41:05
Air Blank	0.000	10:41:35

Sarah Cione
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

6 of 7 SC

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TSY 4-12-2010

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department

DATE 4/1/10 TIME 1025

Intoxilyzer Serial # 80-000331 Location vehicle #432

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.102 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7 sc

154

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 88-002331
04/01/2010 10:25:41

Standard Lot#: 654682
Last Changed By: C. CONLEY #34517

OPS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:26:06
Cal Check	0.102	10:26:27
Air Blank	0.000	10:26:57

Sarah Cione
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.