



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TSy
1.7.2013

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000331 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 12/5/12 Cal 1 0.100 AC Cal 2 0.100 AC

Between the dates of - and - the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.

The instrument was returned to service on 12/5/12.

Sarah Cione

Sarah L. Cione


12/17/12

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 17th, day of December, 20 12.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 12
sc

INTOXILYZER SERVICE RECORD

TS4
1-7-2013

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-000331

Date Returned to Service: 12/5/12

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

2 of 12 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY

DS

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

1.7.2013

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 12/5/12 TIME 1631

INTOXILYZER SERIAL # 80-000331 LOCATION Crime Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.

(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.

(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed

(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.

(✓) 4. Standard Calibration Check standard 0. 0.100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

INTOXILYZER 8000
Location: TUSCON PD
Serial Number: 80-000331
Core Version: 8105.48
12/05/2012

16:32:38

3 of 12 sc
B4

Standard Lot#: 13712100A2
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:33:24
Diagnostic Test	Pass	16:33:58
Air Blank	0.000	16:34:25
0.100 Cal Check	0.100	16:34:46
Air Blank	0.000	16:35:16
Subject Test	0.000	16:35:32
Air Blank	0.000	16:36:00
Wait	ABT*	
Air Blank	0.000	16:36:35

*Sequence Aborted

Not a Successfully Completed Test Sequence

4 of 12_{sc}
Bh

INTOXILYZER 8000
Location: TUSCON PD
Serial Number: 80-000331
Core Version: 8105.48
12/05/2012 16:36:42

Standard Lot#: 13712100A2
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:37:28
Diagnostic Test	Pass	16:38:02
Air Blank	0.000	16:38:29
0.100 Cal Check	0.101	16:38:50
Air Blank	0.000	16:39:20
Subject Test	RFI*	16:39:35
Air Blank	IPS**	16:39:43

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

5 of 12 sc
134

INTOXILYZER 8000
Location: TUSCON PD
Serial Number: 80-000331
Core Version: 8105.48
12/05/2012 . 16:39:55

Standard Lot#: 13712100A2
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:40:38
Diagnostic Test	Pass	16:41:12
Air Blank	0.000	16:41:39
0.100 Cal Check	0.101	16:42:00
Air Blank	0.000	16:42:30
Subject Test	0.165*	16:45:37
Air Blank	0.000	16:46:10
Air Blank	0.000	16:46:37
Subject Test	XXX**	16:47:00
Air Blank	0.000	16:47:35

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUSCON PD
Serial Number: 80-000331
Core Version: 8105.48
12/05/2012 16:47:48

6 of 12 sc

Sh

Standard Lot#: 13712100A2
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	16:48:18
0.100 Cal Check	0.100	16:48:39
Air Blank	0.000	16:49:09

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

7 of 12 sc

TS4
1-7-2013

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 12/5/12 TIME 1631
Intoxilyzer Serial # 80-000331 Location Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000
Location: TUSCON PD
Serial Number: 80-000331
Core Version: 8105.48
12/05/2012 16:31:00

8 of 12_{sc}
B4

Standard Lot#: 13712100A2
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	16:31:29
0.100 Cal Check	0.100	16:31:50
Air Blank	0.000	16:32:19

Preventative Maintenance
Performed -- Timer Reset



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

9 of 12_{sc}

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

Date: 12/5/12
QAS: Sarah Cione
Intoxilyzer Serial Number: 80-000331
New Lot Number: 13712100A2
New Expiration Date: 7/1/14

TJ4
1.7.2013

- (✓) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 2. Remove the old gas tank and install the new one.
- (✓) 3. Hit the "ESC" key twice and swipe the QAS card.
- (✓) 4. Go to the "M-Maintenance" menu and select "S-Standard Change." Select "D" for dry gas and enter the following information:
 Enter Target Value: 0.100
 Enter Control Lot #: From the tank label
 Enter Exp Date: From the tank label (mm/dd/yyyy)
 Edit info Y/N? Y or N
 Update Std Info? Y ENTER
 "Please wait saving..."
- (✓) 5. Hit the "ESC" key twice.
- (✓) 6. Go to the "M-Maintenance" menu and select "T-Control Testing." Select "D" for dry control test. After selection is made press ENTER.
- (✓) 7. Air blank completed.
- (✓) 8. Calibration check completed. Test results 0. 101 AC.
- (✓) 9. Air blank completed.
- (✓) 10. Remove ~~and sign~~ printed record. Attach the record to the completed checklist.
_{sc}

SIGNATURE Sarah Cione

10 of 12_{sc}
TJh

INTOXILYZER 8000
Location: TUSCON PD
Serial Number: 80-000331
Core Version: 8105.48
12/05/2012 16:29:27

Standard Lot#: 13712100A2
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	16:29:55
0.100 Cal Check	0.101	16:30:16
Air Blank	0.000	16:30:46

Preventative Maintenance
Performed -- Timer Reset

12 of 12 sc

TS4

Certificate of Calibration

This is to certify the calibration of **Intoxilyzer**® serial number 80-000331, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58 No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date Nov. 16, 2012 Signed Doh Polier

Technician

EMI INC.

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650517 Rev. A

Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.