



CERTIFICATION by CUSTODIAN of RECORDS  
Tucson Police Department Crime Laboratory

1 of 2 sc  
B4  
2-25-08

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-000332 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: N/A Cal 1 N/A AC Cal 2 N/A AC

Between the dates of \_\_\_\_\_ and \_\_\_\_\_ the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 12/28/07.  
The instrument was returned to service on \_\_\_\_\_.


Sarah Cione

Sarah L. Cione

2/12/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 12, day of February, 2008.



Kathy A Kinney  
Notary Public

\*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT  
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

2 of 2 SC  
DS

Date: 12/28/07

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-000332

Reason For Service: Date on front panel was 18/29/2087. Time/Clock  
battery needs to be replaced.

Date Out of Service: 12/28/07

Calibration Check completed prior to removal from service

Date: could not complete

SQAP completed prior to removal from service

Date: could not complete

Corrective Action:  Sent to CMI, Inc. (1/17/08)

COBRA data downloaded prior to shipping



1 of 15C  
B4

### Return Material Authorization Form

This form **MUST** be completed and enclosed with item(s) being returned.  
*Failure to compete and return this form may cause delays in crediting / repairing.*

1. Item(s) Returned:  Instrument  Supplies  Other  
(Note: please ship items in their original shipping container(s) or a similar protective box.)

2. Instrument Model: Intox 8000 Serial Number 80-000332

3. Bill To Address: Finance / Accounts Payable Ship To Address: (NO PO BOXES) Tucson Police Department  
P.O. Box 27450 Crime Lab  
TUCSON, AZ 85726-7450 270 S. Stone Ave  
TUCSON, AZ 85701-1917

Customer # 105075  
Name Terry Gallegos TPD Crime Lab  
Phone: (520) 791-4494 Fax: (520) 791-4642  
Email: Terry.Gallegos@TucsonAZ.gov

4. Reason for Return: Note - If you are sending an item for repair, please give a detailed description of the problem. Please list any special instructions that you may have concerning this return. Note - Returns for credit may be subject to a 15% restocking charge.  
Time/Clock battery needs to be replaced. Clip holding battery in place is broken as well.

5. Purchase Order Number (attach a copy of P.O. if applicable) 20669

6. I Authorize All Repairs:  Yes  No

7. I Authorize Repairs Up To: \$ \_\_\_\_\_

8. I Require An Estimate Regardless of Cost  Yes  No

**Note: An evaluation fee will apply to estimates that are not repaired.**

9. Authorized By:  
Terry Gallegos  
Name (Please Print)  
[Signature]  
Signature Date 1.17.2008

Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.