



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

*By
6-9-09*

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000332 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 6/5/09 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 5/7/09 and 6/5/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione


6/9/09

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this *9th*, day of *June*, 20*09*.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

1 of 7 sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

TBY
6.9.09

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department
DATE 6/5/09 TIME 11:34
INTOXILYZER SERIAL # 80-000332 LOCATION C12 - Vehicle #945

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7sc

134

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
06/05/2009 11:36:25

Standard Lot#: 8112011
Last Changed By: S. CIONE #50616

OAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SDAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:37:12
Diagnostic Test	Pass	11:37:41
Air Blank	0.000	11:38:08
0.100 Cal Check	0.098	11:38:27
Air Blank	0.000	11:38:57
Subject Test	0.000	11:39:11
Air Blank	0.000	11:39:39
Five Minute Wait	ABT*	
Air Blank	0.000	11:40:12

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

3 of 7_{sc}

134

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
06/05/2009 11:40:19

Standard Lot#: 8112011
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	11:41:00
Air Blank	IPS**	11:41:08

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

4 of 7 sc
TSH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
06/05/2009 11:41:16

Standard Lot#: 8112011
Last Changed By: S. CIONE #50616

OAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:41:59
Diagnostic Test	Pass	11:42:28
Air Blank	0.000	11:42:55
0.100 Cal Check	0.098	11:43:14
Air Blank	0.000	11:43:44
Subject Test	0.047*	11:46:52
Air Blank	0.000	11:47:24
Air Blank	0.000	11:47:52
Subject Test	XXX**	11:48:15
Air Blank	0.000	11:48:49

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7_{sc}

134

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
06/05/2009 11:49:13

Standard Lot#: 8112011
Last Changed By: S. CIONE #50616

OAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:49:40
Cal Check	0.098	11:49:59
Air Blank	0.000	11:50:28

Sarah Cione
Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6 of 7 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

BY
6-9-09

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 6/5/09 TIME 11:34
Intoxilyzer Serial # 80-000332 Location C12-Vehicle#945

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7sc

154

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
06/05/2009 11:34:57

Standard Lot#: 8112011
Last Changed By: S. CIONE #50616

OAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:35:24
Cal Check	0.099	11:35:43
Air Blank	0.000	11:36:13

Sarah Cione
Operator's Signature
