CERTIFICATION by CUSTODIAN of RECORDS

*Tucson Police Department Crime Laboratory*
*270 S. Stone Avenue Tucson, AZ 85701-1917*

TCH 7.28.09

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000332 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 07/23/09 Cal 1 0.097 AC Cal 2 0.097 AC

Between the dates of 06/25/09 and 07/23/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Cheryl K. Straub

Cheryl K. Straub

07/28/09

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 28, day of July, 2009.

A.M. Hernández

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TSH 7-28-09

QA SPECIALIST Cheryl Straub AGENCY Tucson Police Department

DATE 07/23/09 TIME 7:56

INTOXILYZER SERIAL # 80-000332 LOCATION TEAM II/DUI

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

( ✓ ) 1. Clock time check.
( ✓ ) 2. Date check.

OPERATIONAL TESTS

( ✓ ) 1. Alcohol-free subject test result 0. 000 AC.
( ✓ ) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
( ✓ ) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
( ✓ ) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS Deficient sample test missed on 07/23/09
This test was performed on 07/27/09.
Duplicate tests were performed because of
external printer failure - one test interrupted
by RFI. cs

SIGNATURE Cheryl Straub

DPS Form Exh G-4 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

2 of 10 CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
07/23/2009 07:59:09

Standard Lot#: 811201I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:00:18 |
| Diagnostic Test | Pass | 08:00:45 |
| Air Blank | 0.000 | 08:01:13 |
| 0.100 Cal Check | 0.097 | 08:01:31 |
| Air Blank | 0.000 | 08:02:01 |
| Subject Test | 0.000 | 08:02:31 |
| Air Blank | ABT* | 08:02:39 |
| Air Blank | 0.000 | 08:03:07 |

*Sequence Aborted

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3710 CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
07/23/2009 08:04:02

Standard Lot#: 811201I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | RFI* | 08:04:56 |
| Air Blank | IPS** | 08:05:16 |

*RFI Detect
**Improper Sample

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4810
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
07/27/2009 12:34:56

Standard Lot#: 811201I
Last Changed By: S. CIONE #50616

54

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: / /
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 12:35:54 |
| Diagnostic Test | Pass | 12:36:22 |
| Air Blank | 0.000 | 12:36:49 |
| 0.100 Cal Check | 0.097 | 12:37:08 |
| Air Blank | 0.000 | 12:37:38 |
| Subject Test | 0.115* | 12:40:46 |
| Air Blank | 0.000 | 12:41:18 |
| Air Blank | 0.000 | 12:41:46 |
| Subject Test | 0.000** | 12:41:57 |
| Air Blank | 0.000 | 12:42:26 |

*Deficient Sample
**Sequence Aborted

**Not a Successfully Completed Test Sequence**

Reprinted From Database

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 9 10
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
07/27/2009 12:53:07

Standard Lot#: 811201I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: / /
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 12:54:04 |
| Diagnostic Test | Pass | 12:54:31 |
| Air Blank | 0.000 | 12:54:58 |
| 0.100 Cal Check | 0.097 | 12:55:17 |
| Air Blank | 0.000 | 12:55:47 |
| Subject Test | 0.000* | 12:56:01 |
| Air Blank | RFI** | 12:56:15 |

*RFI Detect
**RFI Detect

**Not a Successfully Completed Test Sequence**

Reprinted From Database

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6 of 10 CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
07/27/2009 12:57:09

Standard Lot#: 811201I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: / /
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 12:58:13 |
| Diagnostic Test | Pass | 12:58:41 |
| Air Blank | 0.000 | 12:59:08 |
| 0.100 Cal Check | 0.097 | 12:59:27 |
| Air Blank | 0.000 | 12:59:56 |
| Subject Test | 0.121* | 13:03:05 |
| Air Blank | 0.000 | 13:03:37 |
| Air Blank | 0.000 | 13:04:05 |
| Subject Test | 0.030** | 13:07:13 |
| Air Blank | 0.000 | 13:07:46 |

*Deficient Sample
**Deficient Sample

**Not a Successfully Completed Test Sequence**

Reprinted From Database

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 8 10
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
07/23/2009 08:09:43

104

Standard Lot#: 811201I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:10:47 |
| Diagnostic Test | Pass | 08:11:15 |
| Air Blank | 0.000 | 08:11:42 |
| 0.100 Cal Check | 0.098 | 08:12:01 |
| Air Blank | 0.000 | 08:12:30 |
| Subject Test | XXX* | 08:15:31 |
| Air Blank | 0.000 | 08:16:05 |

*Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

8 910
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
07/23/2009 08:16:46

---

Standard Lot#: 811201I
Last Changed By: S. CIONE #50616

---

QAS: C. STRAUB #50772
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:17:13 |
| Cal Check | 0.097 | 08:17:32 |
| Air Blank | 0.000 | 08:18:02 |

Cheyl Straub
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

9810 CS

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TJH 7.28.09

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department

DATE 07/23/09 TIME 07:56

Intoxilyzer Serial # 80-000332 Location TEAM II / DUI

( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

( ✓ ) 4. Air blank completed.

( ✓ ) 5. Calibration check completed. Test results 0.097 AC.

( ✓ ) 6. Air blank completed.

( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

10 9 10
CS

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
07/23/2009 07:56:14

---

Standard Lot#: 811201I
Last Changed By: S. CIONE #50616

---

QAS: C. STRAUB #50772
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 07:56:41 |
| Cal Check | 0.097 | 07:57:00 |
| Air Blank | 0.000 | 07:57:30 |

Cheryl Straub
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.