CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000332 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TJ 11.17.09

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11-12-09 Cal 1 0.097 AC Cal 2 0.096 AC

Between the dates of 10-15-09 and 11-12-09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

11-17-09

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 17, day of November, 2009.

Kim Messinger

Notary Public

OFFICIAL SEAL
KIM MESSINGER
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. June 20, 2010

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7
Cu

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

TBG
11-17-09

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 11-12-09 TIME 1306

INTOXILYZER SERIAL # 80-000332 LOCATION ODW DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 096 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7
CW

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
11/12/2009 13:09:01

Standard Lot#: 801002I
Last Changed By: W. HONOMICHL #42992

BY

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:09:47
Diagnostic Test	Pass	13:10:15
Air Blank	0.000	13:10:42
0.100 Cal Check	0.097	13:11:01
Air Blank	0.000	13:11:30
Subject Test	0.000	13:12:03
Air Blank	ABT*	13:12:10
Air Blank	0.000	13:12:38

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
11/12/2009 13:12:57

3 of 7
Cu

Standard Lot#: 801002I
Last Changed By: W. HONOMICHL #42992

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	13:13:30
Air Blank	IPS**	13:13:40

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4567
Cw

BH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-000332
11/12/2009 13:14:27

Standard Lot#: 801002I
Last Changed By: W. HONOMICHL #42992

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:15:38
Diagnostic Test	Pass	13:16:05
Air Blank	0.000	13:16:33
0.100 Cal Check	0.098	13:16:51
Air Blank	0.000	13:17:21
Subject Test	0.346*	13:20:29
Air Blank	0.000	13:21:02
Air Blank	0.000	13:21:29
Subject Test	XXX**	13:21:50
Air Blank	0.000	13:22:24

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
11/12/2009 13:22:46

5 of 7
Cr

Standard Lot#: 801002I
Last Changed By: W. HONOMICHL #42992

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:23:13
Cal Check	0.096	13:23:32
Air Blank	0.000	13:24:01

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

6 of 7

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

Cu

STANDARD CALIBRATION CHECK PROCEDURE

TSH 11-17-09

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 11-12-09 TIME 1306

Intoxilyzer Serial # 80-000332 Location ODW DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000332
11/12/2009 13:06:41

Standard Lot#: 801002I
Last Changed By: W. HONOMICHL #42992

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:07:08
Cal Check	0.097	13:07:27
Air Blank	0.000	13:07:57

7567
cu

TBG

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.