



CERTIFICATION by CUSTODIAN of RECORDS  
 Tucson Police Department Crime Laboratory  
 1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-000332 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

*TS4  
2-23-2012*

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 2/1/12 Cal 1 - AC Cal 2 - AC

Between the dates of - and - the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 2/1/12.  
 The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione


2/21/12

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 21<sup>st</sup>, day of February, 20 12.

Elma Escarcega

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.



TUCSON POLICE DEPARTMENT  
CRIME LABORATORY

1 of 2 sc

INTOXILYZER SERVICE RECORD

TJY  
2.23.2012

Date: 2/1/12

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-000332

Reason For Service: Downloaded software version 8105.48  
then intox would not turn back on

Date Out of Service: 2/1/12

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action:  Sent to CMI, Inc.  
per Becky Taylor, a chip needs to be replaced to update date on intox

Attn: Becky Taylor

2 of 2 sc

**EMI INC.**  
**Service Evaluation Form**

This form **MUST** be completed and enclosed with instrument to be serviced.  
Failure to complete and return this form may cause delays in service.

154

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number 105075 (contact Customer Service)  
Name Terry Gallegos Phone: (520) 837-7677  
Fax: (520) 791-4642 Email: Terry.Gallegos@tucsonaz.gov

2. Bill to Address: Ship to Address:  
Finance / Accounts Payable Tucson Police Department Crime Lab  
Po Box 27450 1306 W. Miracle Mile  
Tucson, AZ 85726-7450 Tucson, AZ 85705

3. Serial Number: 80-000332 Instrument Model: I8000

4. Detailed Description of Problem:  
downloaded new software version on 2/1/2012. Won't turn back on.

\*\*\*Hazardous Material Warning! - DO NOT return gas cylinders with instrument!\*\*\*

I Authorize Repairs Up To:  All  \$250  \$500  \$750  Other \$ \_\_\_\_\_  
Purchase Order Number (attach a copy of P.O. if applicable) 34770

Authorized By:

Terry Gallegos Toxicology Coordinator  
Name (Please Print) Title

[Signature] 2.1.2012  
Signature Date

Ship to:

CMI, Inc.  
Attn: Service Dept.  
316 East Ninth Street  
Owensboro, KY 42303

No, please send estimate before repairs are made.  
Note: An estimate will be faxed before performing any repairs and may cause delays in service.  
An evaluation fee (\$79.00 or actual costs) will apply to estimates that are not repaired.

316 EAST NINTH STREET OWENSBORO, KENTUCKY

Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.