



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 1306 W. Miracle Mile, Tucson, AZ 85705

TSy
 6-27-2013

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000334 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 6/20/13 Cal 1 0.102 AC Cal 2 0.100 AC

Between the dates of 5/23/13 and 6/20/13 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
 The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione

6/25/13


ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 25th, day of June, 2013.

Elma Escarcega

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5



Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7 sc
TSh
6.27.2013

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 6/20/13 TIME 936

INTOXILYZER SERIAL # 80-000334 LOCATION DUIHQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
06/20/2013

09:39:23

2 of 7 sc
BY

Standard Lot#: 09312100A1
Last Changed By: W. HONOMICHL #42992

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:40:14
Diagnostic Test	Pass	09:40:48
Air Blank	0.000	09:41:16
0.100 Cal Check	0.101	09:41:37
Air Blank	0.000	09:42:06
Subject Test	0.000	09:42:21
Air Blank	0.000	09:42:48
Wait	ABT*	
Air Blank	0.000	09:43:22

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
06/20/2013 09:47:08

3 of 7 sc
Tm

Standard Lot#: 09312100A1
Last Changed By: W. HONOMICHL #42992

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:48:05
Diagnostic Test	Pass	09:48:39
Air Blank	0.000	09:49:06
0.100 Cal Check	0.100	09:49:27
Air Blank	0.000	09:49:57
Subject Test	RFI*	09:50:12
Air Blank	IPS**	09:50:22

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
06/20/2013 09:50:28

4 of 7 sc
T24

Standard Lot#: 09312100A1
Last Changed By: W. HONOMICHL #42992

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:51:24
Diagnostic Test	Pass	09:51:58
Air Blank	0.000	09:52:26
0.100 Cal Check	0.100	09:52:47
Air Blank	0.000	09:53:16
Subject Test	0.094*	09:56:24
Air Blank	0.000	09:56:56
Air Blank	0.000	09:57:23
Subject Test	XXX**	09:58:12
Air Blank	0.000	09:58:46

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
06/20/2013 09:59:06

5 of 7 sc
BH

Standard Lot#: 09312100A1
Last Changed By: W. HONOMICHL #42992

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:59:34
0.100 Cal Check	0.100	09:59:55
Air Blank	0.000	10:00:24

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7 sc
Dy
6.27.2013

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 6/20/13 TIME 936
Intoxilyzer Serial # 80-000334 Location DUIHQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 102 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
06/20/2013 09:36:57

7 of 7 sc
T&G

Standard Lot#: 09312100A1
Last Changed By: W. HONOMICHL #42992

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:37:26
0.100 Cal Check	0.102	09:37:47
Air Blank	0.000	09:38:16

Preventative Maintenance
Performed -- Timer Reset