



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

SC
9/9/14

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000334 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 8-21-2014 Cal 1 0.098 AC Cal 2 0.096 AC

Between the dates of 7-24-2014 and 8-21-2014 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.


Terry Gallegos

Terry S. Gallegos

9-8-2014

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 8th, day of September, 2014.



Elma Escarcega

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
TS4
sc
9/9/14

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 8.21.2014 TIME 1142

INTOXILYZER SERIAL # 80-000334 LOCATION ODM Sgt's Office

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 096 AC

Instrument is operating properly and accurately. YES NO

COMMENTS _____

SIGNATURE Terry Gallegos

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7
TSC
3C

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
08/21/2014 11:44:31

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

OAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SOAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:45:26
Diagnostic Test	Pass	11:46:02
Air Blank	0.000	11:46:30
0.100 Cal Check	0.097	11:46:51
Air Blank	0.000	11:47:20
Subject Test	0.000	11:47:34
Air Blank	0.000	11:48:01
Wait	ABT*	
Air Blank	0.000	11:49:12

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

3 of 7
TS4

5C

INTOX:LYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
08/21/2014 11:49:31

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST,SOAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:50:23
Diagnostic Test	Pass	11:50:59
Air Blank	0.000	11:51:26
0.100 Cal Check	RFI*	11:51:45
Air Blank	IPS**	11:52:02

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

4 of 7
TS4

52

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
08/21/2014 11:52:23

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

OAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:53:14
Diagnostic Test	Pass	11:53:51
Air Blank	0.000	11:54:18
0.100 Cal Check	0.098	11:54:39
Air Blank	0.000	11:55:08
Subject Test	0.431*	11:58:16
Air Blank	0.000	11:58:50
Air Blank	0.000	11:59:18
Subject Test	XXX**	11:59:32
Air Blank	0.000	12:00:05

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7
TSh

SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-00334
Core Version: 8105.48
08/21/2014 12:00:28

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

QAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	12:00:58
0.100 Cal Check	0.096	12:01:18
Air Blank	0.000	12:01:48

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

leaf 7
TS4

sc
9/9/14

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 8-21-2014 TIME 1142
Intoxilyzer Serial # 80-000334 Location DDM Sgt's Office

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7
TS4
SC

INTOXILYZER 8000
Instrument Initialization
11:01:26 08/21/2014

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
08/21/2014 11:42:54

Standard Lot#: 25913100A2
Last Changed By: S. CIONE #50616

OAS: T. GALLEGOS #28847
TUCSON PD

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	11:43:22
0.100 Cal Check	0.098	11:43:42
Air Blank	0.000	11:44:12