



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

T54
10-29-2014

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000334 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10/14/14 Cal 1 - AC Cal 2 - AC

Between the dates of - and - the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 10/14/14.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione

10/20/14


ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 20th, day of October, 2014.

Elma Escarcega

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5



Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

1 of 2
TS
10-29-2014

Date: 10/14/14

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-000334

Reason For Service: Notified by Sgt. Ohara that instrument would not go to Ready Mode

Date Out of Service: 10/14/14

Calibration Check completed prior to removal from service

Date: —

SQAP completed prior to removal from service

Date: —

Corrective Action: Sent to CMI, Inc.



Service Evaluation Form

2 of 2 sc
B4

This form MUST be completed and enclosed with instrument to be serviced.
Failure to complete and return this form may cause delays in service.

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number 105075 (contact Customer Service)
Name Terry Gallegos Phone: (520) 837-7677
Fax: (520) 791-4642 Email: Terry.Gallegos@tucsonaz.gov

2. Bill to Address: Finance / Accounts Payable Ship to Address: Tucson Police Department Crime Lab
Po Box 27450 1306 W. Miracle Mile
Tucson, AZ 85726-7450 Tucson, AZ 85705

3. Serial Number: 80-000334 Instrument Model: I 8000

4. Detailed Description of Problem: Will not come out of "Not Ready Mode". Sample Cell not heating up.
Both keyboard hinges broken

Hazardous Material Warning! - DO NOT return gas cylinders with instrument!

I Authorize Repairs Up To: All \$250 \$500 \$750 Other \$ _____
Purchase Order Number (attach a copy of P.O. if applicable) 34770

Authorized By:
Terry Gallegos Toxicology Coordinator
Name (Please Print) Title
Terry Gallegos 10.17.2014
Signature Date

Ship to:
CMI, Inc.
Attn: Service Dept.
316 East Ninth Street
Owensboro, KY 42303

No, please send estimate before repairs are made.
Note: An estimate will be faxed before performing any repairs and may cause delays in service.
An evaluation fee (\$79.00 or actual costs) will apply to estimates that are not repaired.