



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000334 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TSy
7-6-2015

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5/28/2015 Cal 1 0.097 AC Cal 2 0.096 AC

Between the dates of 4/30/2015 and 5/28/2015 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on .
The instrument was returned to service on .

William L Loomis

William L. Loomis


6/8/2015

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 8th, day of June, 20 15.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
WLL
TS4 7-6-2015

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department

DATE 5/28/2015 TIME 08:47

INTOXILYZER SERIAL # 80-000334 LOCATION ODM - SGT'S OFFICE

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 096 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Will Loomis

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 8C-000334
Core Version: 8105.48
05/28/2015 08:54:53

2 of 7
will
124

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:56:29
Diagnostic Test	Pass	08:57:06
Air Blank	0.000	08:57:33
0.100 Cal Check	0.097	08:57:54
Air Blank	0.000	08:58:23
Subject Test	0.000	08:58:40
Air Blank	0.000	08:59:17
Wait	ABT*	
Air Blank	0.000	08:59:43

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
05/28/2015 09:01:47

3 of 7
WLL
154

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #26847

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:03:14
Diagnostic Test	Pass	09:03:51
Air Blank	0.000	09:04:18
0.100 Cal Check	RFI*	09:04:34
Air Blank	IPS**	09:04:46

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

4 of 7
wll
TJH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
05/28/2015 09:06:43

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

DAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:08:21
Diagnostic Test	Pass	09:08:57
Air Blank	0.000	09:09:25
0.100 Cal Check	0.098	09:09:46
Air Blank	0.000	09:10:15
Subject Test	0.139*	09:13:23
Air Blank	0.000	09:13:56
Air Blank	0.000	09:14:23
Subject Test	XXX**	09:14:40
Air Blank	0.000	09:15:15

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7
wll
B4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
05/28/2015 09:19:46

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:20:16
0.100 Cal Check	0.096	09:20:37
Air Blank	0.000	09:21:07

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
WLL
TS4 7.6.2015

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 5/28/2015 TIME 08:47
Intoxilyzer Serial # 80-000334 Location ODM - SGT'S OFFICE

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Will Loomis

DPS Form Exh G-2 (Rev. 05-01)

7 of 7
WLL
TSh

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
05/28/2015 08:47:15

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:47:45
0.100 Cal Check	0.097	08:48:06
Air Blank	0.000	08:48:36

Preventative Maintenance
Performed -- Timer Reset