



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000334 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TS4
8.6.2015

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7/23/2015 Cal 1 0.096 AC Cal 2 0.095 AC

Between the dates of 6/25/2015 and 7/23/2015 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on _____.
The instrument was returned to service on _____.

William L Loomis

William L. Loomis


7/27/2015

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 27th, day of July, 2015.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

1 of 7
WU

STANDARD QUALITY ASSURANCE PROCEDURE

TSh
8.6.2015

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department

DATE 7/23/2015 TIME 09:18

INTOXILYZER SERIAL # 80-000334 LOCATION ODM'SGT'S OFFICE

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 095 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Will Loomis

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-00334
Core Version: 8105.48
07/23/2015 09:21:49

2 of 7
WLV
TS4

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:22:45
Diagnostic Test	Pass	09:23:21
Air Blank	0.000	09:23:49
0.100 Cal Check	0.096	09:24:10
Air Blank	0.000	09:24:39
Subject Test	0.000	09:25:02
Air Blank	0.000	09:25:30
Wait	ABT*	
Air Blank	0.000	09:26:05

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

3 of 7
wlv
TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
07/23/2015 09:27:53

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:28:46
Diagnostic Test	Pass	09:29:22
Air Blank	0.000	09:29:50
0.100 Cal Check	RFI*	09:30:04
Air Blank	IPS**	09:30:13

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-010334
Core Version: 8155.48
07/23/2015 09:31:38

4 of 7
www
TS4

Standard Lot#: 11514100A2
Last Charged By: T. GALLEGOS #28847

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:32:38
Diagnostic Test	Pass	09:33:15
Air Blank	0.000	09:33:42
0.100 Cal Check	0.096	09:34:03
Air Blank	0.000	09:34:33
Subject Test	0.179*	09:37:40
Air Blank	0.000	09:38:12
Air Blank	0.000	09:38:39
Subject Test	XXX**	09:39:02
Air Blank	0.000	09:39:36

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully
Completed
Test Sequence

5 of 7
wll
TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-00334
Core Version: 8105.48
07/23/2015 09:40:37

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:41:07
0.100 Cal Check	0.095	09:41:28
Air Blank	0.000	09:41:58

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
WVL

TSK
8/6/2015

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 7/23/2015 TIME 09:18
Intoxilyzer Serial # 80-000334 Location ODM-SGT'S OFFICE

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 096 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Will Loomis

DPS Form Exh G-2 (Rev. 05-01)

7 of 7
TS4
WU

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
07/23/2015 09:18:24

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:18:56
0.100 Cal Check	0.096	09:19:17
Air Blank	0.000	09:19:46

Preventative Maintenance
Performed -- Timer Reset