



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TS4
5.17.2016

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000334 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4-29-16 Cal 1 0.100 AC Cal 2 0.098 AC

Between the dates of — and — the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on 4-22-16.

[Signature]
Amy L. Autret


5-3-16

ADPS QAS Permit # 33718

SUBSCRIBED AND SWORN to before me this 3rd, day of May, 2016.

[Signature]

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT
CRIME LABORATORY
INTOXILYZER SERVICE RECORD

1 of 14
JLB

TS4
5-17-2016

QAS: Amy Autret 53546

Intoxilyzer Serial Number: 80-000334

Date Returned to Service: 4-22-16

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

2 of 14
TJG
5.17.2016

QA SPECIALIST Amy Autret 53546 AGENCY Tucson Police Department

DATE 4-29-16 TIME 1348

INTOXILYZER SERIAL # 80-000334 LOCATION Crime Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.


OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE



INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
04/29/2016

13:54:20

3 of 14
Old
TS4

Standard Lot#: 06616100A1
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:55:18
Diagnostic Test	Pass	13:55:52
Air Blank	0.000	13:56:20
0.100 Cal Check	0.098	13:56:40
Air Blank	0.000	13:57:10
Subject Test	0.000	13:57:44
Air Blank	0.000	13:58:12
Wait	ABT*	
Air Blank	0.000	13:58:46

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
04/29/2016

13:59:03

4 of 14
old
104

Standard Lot#: 06616100A1
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:00:05
Diagnostic Test	Pass	14:00:39
Air Blank	0.000	14:01:06
0.100 Cal Check	RFI*	14:01:20
Air Blank	IPS**	14:01:30

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
04/29/2016

14:01:40

5 of 14
JL
TS4

Standard Lot#: 06616100A1
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:02:39
Diagnostic Test	Pass	14:03:13
Air Blank	0.000	14:03:41
0.100 Cal Check	0.099	14:04:02
Air Blank	0.000	14:04:31
Subject Test	0.300*	14:07:39
Air Blank	0.000	14:08:12
Air Blank	0.000	14:08:39
Subject Test	XXX**	14:08:59
Air Blank	0.000	14:09:33

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
04/29/2016 14:11:22

6 of 14
JLS
T34

Standard Lot#: 06616100A1
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	14:11:51
0.100 Cal Check	0.098	14:12:12
Air Blank	0.000	14:12:42

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

2 of 14
ADD
154

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Amy Autret 53546 AGENCY Tucson Police Department
DATE 4-29-16 TIME 1348
Intoxilyzer Serial # 80-000334 Location Crime Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____



DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
04/29/2016 13:48:30

SOFM
JL
134

Standard Lot#: 11514100A2
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:49:00
0.100 Cal Check	0.100	13:49:21
Air Blank	0.000	13:49:50

Preventative Maintenance
Performed -- Timer Reset



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

9 of 14
TJG
5.17.2016

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

Date: 4-29-16

QAS: Amy Autret 53546

Intoxilyzer Serial Number: 80-000334

New Lot Number: 06616100A1

New Expiration Date: 4-5-2018

- (✓) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 2. Remove the old gas tank and install the new one.
- (✓) 3. Hit the "ESC" key twice and swipe the QAS card.
- (✓) 4. Go to the "M-Maintenance" menu and select "S-Standard Change." Select "D" for dry gas and enter the following information:
 Enter Target Value: 0.100
 Enter Control Lot #: From the tank label
 Enter Exp Date: From the tank label (mm/dd/yyyy)
 Edit info Y/N? Y or N
 Update Std Info? Y ENTER
 "Please wait saving..."
- (✓) 5. Hit the "ESC" key twice.
- (✓) 6. Go to the "M-Maintenance" menu and select "T-Control Testing." Select "D" for dry control test. After selection is made press ENTER.
- (✓) 7. Air blank completed.
- (✓) 8. Calibration check completed. Test results 0. 099 AC.
- (✓) 9. Air blank completed.
- (✓) 10. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
04/29/2016 13:52:35

10 of 14
TS4

Standard Lot#: 06616100A1
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:53:05
0.100 Cal Check	0.099	13:53:26
Air Blank	0.000	13:53:55

Preventative Maintenance
Performed -- Timer Reset



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

11 of 14
TSJ
5.17.2016

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

Date: 4-22-2016

QAS: Amy Antret 53546

Intoxilyzer Serial Number: 80-000334

New Lot Number: 11514100A2

New Expiration Date: 5/1/2016

- (✓) ✓ 1. Intoxilyzer 8000 display reads ^{contact QAS} ~~"PUSH BUTTON TO START"~~ *all*
- (✓) ✓ 2. Remove the old gas tank and install the new one.
- (✓) ✓ 3. Hit the "ESC" key twice and swipe the QAS card.
- (✓) ✓ 4. Go to the "M-Maintenance" menu and select "S-Standard Change." Select "D" for dry gas and enter the following information:

Enter Target Value: 0.100
 Enter Control Lot #: From the tank label
 Enter Exp Date: From the tank label (mm/dd/yyyy)
 Edit info Y/N? Y or N
 Update Std Info? Y ENTER
 "Please wait saving..."

- (✓) ✓ 5. Hit the "ESC" key twice.
- (✓) ✓ 6. Go to the "M-Maintenance" menu and select "T-Control Testing." Select "D" for dry control test. After selection is made press ENTER.
- (✓) ✓ 7. Air blank completed.
- (✓) ✓ 8. Calibration check completed. Test results 0. 000 AC. 0.099 AC
- (✓) ✓ 9. Air blank completed.
- (✓) ✓ 10. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

120414
TS4
JLB

INTOXILYZER 8000
Instrument Initialization
15:10:30 04/22/2016

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
04/22/2016 15:48:04

Standard Lot#: 11514100A2
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	15:48:39
0.100 Cal Check	0.000*	15:48:59
Air Blank	0.000	15:49:32

*Calibration Check Out of Tolerance

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000334
Core Version: 8105.48
04/22/2016 15:57:28

Standard Lot#: 11514100A2
Last Changed By: A. AUTRET #53546

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	15:59:10
0.100 Cal Check	0.099	15:59:31
Air Blank	0.000	16:00:00

Preventative Maintenance
Performed -- Timer Reset

Gas Tank empty.
Attached new tank
and repeated cal
check.

14 of 14
JAD

TS4

Certificate of Calibration

This is to certify the calibration of **Intoxilyzer**® serial number 80-000334, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58, No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date April 15, 2016 Signed [Signature]
Technician



Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650517 Rev.A