



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
2-4-2009

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000712 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1/18/09 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 12/18/08 and 1/18/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —
The instrument was returned to service on —

Sarah Cione

Sarah L. Cione


1/21/09

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 21st, day of January, 20 09.

Ant Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 8 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

*TSJ
2.4.2009*

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 1/18/09 TIME 19:27

INTOXILYZER SERIAL # 80-000712 LOCATION Vehicle # 866 - Ft. Lowell / Swan

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

2 of 8 sc

TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
01/18/2009 19:29:33

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RF1*	19:30:26
Air Blank	0.000	19:30:54

*RF1 Detect

Not a Successfully
Completed
Test Sequence

3 of 8 sc

SM

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
01/18/2009 19:31:08

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

OAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:32:02
Diagnostic Test	Pass	19:32:31
Air Blank	0.000	19:32:58
0.100 Cal Check	0.100	19:33:17
Air Blank	0.000	19:33:46
Subject Test	0.000	19:34:03
Air Blank	0.000	19:34:30
Five Minute Wait	ABT*	
Air Blank	0.000	19:35:04

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

4 of 8 sc
JBY

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
01/18/2009 19:35:14

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

OAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	19:35:54
Air Blank	IPS**	19:36:05

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

5 of 8_{sc}

PSH

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
01/18/2009 19:36:16

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

OAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:37:06
Diagnostic Test	Pass	19:37:34
Air Blank	0.000	19:38:02
0.100 Cal Check	0.100	19:38:20
Air Blank	0.000	19:38:50
Subject Test	0.049*	19:41:58
Air Blank	0.000	19:42:30
Air Blank	0.000	19:42:57
Subject Test	XXX**	19:43:23
Air Blank	0.000	19:43:56

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

6 of 8 SC
TS4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
01/18/2009 19:44:29

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

OAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	19:44:56
Cal Check	0.100	19:45:15
Air Blank	0.000	19:45:44

Sarah Cione
Operator's Signature

EXHIBIT G-2

7 of 8 sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

TS4
2.4.2009

QA Specialist Sarah Cione AGENCY Tucson Police Department

DATE 1/18/09 TIME 19:27

Intoxilyzer Serial # 80-000712 Location Vehicle # 866 - Ft. Lowell / Swan

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

8 of 8 sc

124

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-000712
01/18/2009 19:27:50

Standard Lot#: 735401G
Last Changed By: S. CLEARY #44524

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	19:28:17
Cal Check	0.101	19:28:36
Air Blank	0.000	19:29:05

Sarah Cione

Operator's Signature