



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 1306 W. Miracle Mile, Tucson, AZ 85705

TS4
 3.12.2015

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-000712 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 3/5/2015 Cal 1 0.099 AC Cal 2 0.097 AC

Between the dates of 2/5/2015 and 3/5/2015 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on _____.
 The instrument was returned to service on _____.

William L Loomis

William L. Loomis


3/11/2015

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 11th, day of March, 20 15.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
WLL
TS4
3.12.2015

QA SPECIALIST Will Loomis 53545 AGENCY Tucson Police Department

DATE 3.5.2015 TIME 09:05

INTOXILYZER SERIAL # 80-000712 LOCATION ODM - SGT'S OFFICE

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Will Loomis

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-000712
Core Version: 8105.48
03/05/2015 09:08:36

2 of 7
WLL
TS4

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:09:31
Diagnostic Test	Pass	09:10:08
Air Blank	0.000	09:10:35
0.100 Cal Check	0.098	09:10:56
Air Blank	0.000	09:11:25
Subject Test	0.000	09:11:44
Air Blank	0.000	09:12:12
Wait	ABT*	
Air Blank	0.000	09:12:47

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-000712
Core Version: 8105.48
03/05/2015 09:14:11

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP, TEST
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:15:09
Diagnostic Test	Pass	09:15:45
Air Blank	0.000	09:16:12
0.100 Cal Check	RF1*	09:16:26
Air Blank	IPS**	09:16:36

*RF1 Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

3 of 7
WLC
TSh

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-000712
Core Version: 8105-48
03/05/2015 09:17:12

4 of 7
WLU
TS4

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SOAP,TEST
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:18:12
Diagnostic Test	Pass	09:18:49
Air Blank	0.000	09:19:16
0.100 Cal Check	0.098	09:19:37
Air Blank	0.000	09:20:06
Subject Test	0.067*	09:23:14
Air Blank	0.000	09:23:46
Air Blank	0.000	09:24:14
Subject Test	XXX**	09:24:30
Air Blank	0.000	09:25:04

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-000712
Core Version: 8105.48
03/05/2015 09:26:59

5 of 7
wrv
TS4

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

GAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:27:29
0.100 Cal Check	0.097	09:27:50
Air Blank	0.000	09:28:19

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
WB
3.12.2015

QA Specialist WILL LOOMIS 53545 AGENCY Tucson Police Department
DATE 3.5.2015 TIME 09:05
Intoxilyzer Serial # 80-000712 Location ODM - SGT'S OFFICE

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Will Loomis

7 of 7
wll
TS4

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-000712
Core Version: 8105.48
03/05/2015 09:05:33

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

OAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:06:05
0.100 Cal Check	0.099	09:06:26
Air Blank	0.000	09:06:55

Preventative Maintenance
Performed -- Timer Reset