



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-000712 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TSL  
7-26-2016

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7/21/2016 Cal 1 0.096 AC Cal 2 0.096 AC

Between the dates of 6/24/2016 and 7/21/2016 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on \_\_\_\_\_.  
The instrument was returned to service on \_\_\_\_\_.

William L Loomis

William L. Loomis


7/21/2016

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 21<sup>st</sup>, day of July, 2016.

Elma Escarcega

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7  
WLL  
TSK 7.26.2016

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department  
DATE 7/21/2016 TIME 13:45  
INTOXILYZER SERIAL # 80-000712 LOCATION CRIME LAB

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ( ✓ ) 1. Clock time check.
- ( ✓ ) 2. Date check.

OPERATIONAL TESTS

- ( ✓ ) 1. Alcohol-free subject test result 0. 000 AC.
- ( ✓ ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- ( ✓ ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- ( ✓ ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 096 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Willie Loomis

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-000712  
Core Version: 8105.48  
07/21/2016

13:48:31

Standard Lot#: 06616100A1  
Last Changed By: A. AUTRET #53546

QAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SQAP, SQAP  
DOB: 11/11/1987  
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:49:59
Diagnostic Test	Pass	13:50:33
Air Blank	0.000	13:51:00
0.100 Cal Check	0.097	13:51:21
Air Blank	0.000	13:51:50
Subject Test	0.000	13:52:12
Air Blank	0.000	13:52:39
Wait	ABT*	
Air Blank	0.000	13:53:15

\*Sequence Aborted

2 of 7  
bwn  
154

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-000712  
Core Version: 8105.48  
07/21/2016 13:54:03

Standard Lot#: 06616100A1  
Last Changed By: A. AUTRET #53546

QAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SQAP, SQAP  
DOB: 11/11/1987  
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:55:33
Diagnostic Test	Pass	13:56:07
Air Blank	0.000	13:56:34
0.100 Cal Check	RFI*	13:56:48
Air Blank	IPS**	13:56:58

\*RFI Detect  
\*\*Improper Sample

3 of 7  
6 WLL  
Tsk

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-000712  
Core Version: 8105.48  
07/21/2016

13:57:21

Standard Lot#: 06616100A1  
Last Changed By: A. AUTRET #53546

QAS: W. LOOMIS #53545  
TUCSON PD

Operator: W. LOOMIS #53545  
TUCSON PD

Subject: SQAP, SQAP

DOB: 11/11/1987

Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:58:12
Diagnostic Test	Pass	13:58:46
Air Blank	0.000	13:59:13
0.100 Cal Check	0.096	13:59:34
Air Blank	0.000	14:00:03
Subject Test	0.056*	14:03:11
Air Blank	0.000	14:03:42
Air Blank	0.000	14:04:09
Subject Test	XXX**	14:04:37
Air Blank	0.000	14:05:10

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

WLL  
3 of 7  
4  
WLL  
134

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-000712  
Core Version: 8105.48  
07/21/2016 14:05:35

-----  
Standard Lot#: 06616100A1  
Last Changed By: A. AUTRET #53546  
-----

QAS: W. LOOMIS #53545  
TUCSON PD  
-----

Operator: W. LOOMIS #53545  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	14:06:06
0.100 Cal Check	0.096	14:06:27
Air Blank	0.000	14:06:56

Preventative Maintenance  
Performed -- Timer Reset

4 of 7  
5 0 WLL  
PG

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

*WLL*  
*6 of 7*  
*6 0 WLL*  
*TR 7-26-2016*

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department  
DATE 7/21/2016 TIME 13:45  
Intoxilyzer Serial # 80-000712 Location CRIME LAB

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 096 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Will Loomis

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-000712  
Core Version: 8105.48  
07/21/2016 13:45:37

-----  
Standard Lot#: 06616100A1  
Last Changed By: A. AUIRET #53546  
-----

QAS: W. LOOMIS #53545  
TUCSON PD  
-----

Operator: W. LOOMIS #53545  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	13:46:07
0.100 Cal Check	0.096	13:46:28
Air Blank	0.000	13:46:58

Preventative Maintenance  
Performed -- Timer Reset

*6 WLL*  
*7 of 7*  
*WLL*  
*TG*  
*7-26-2016*