



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

CU
11-23-07

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001108 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10-26-2007 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 9-26-2007 and 10-26-2007 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on .
The instrument was returned to service on .


Terry S. Gallegos

Terry S. Gallegos

11-21-2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 21, day of November, 2007.



Kathy A Kinney
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

CU
11-23-07
10f8
TS4

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 10-26-2007 TIME 0931

INTOXILYZER SERIAL # 80-001108 LOCATION ODM

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed ✓ ✓
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Terry Gallegos

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number : 80-001108

10/26/2007

09:33:05

cu
2 of 8
B4

Standard Lot#: 719801I

Last Changed By: C. CONLEY

#34517

QAS: T. GALLEGOS

#28847

TUCSON PD CRIME LAB

Operator: T. GALLEGOS

#28847

TUCSON PD CRIME LAB

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|------------------|--------|----------|
| Air Blank | 0.000 | 09:33:52 |
| Diagnostic Test | Pass | 09:34:19 |
| Air Blank | 0.000 | 09:34:46 |
| 0.100 Cal Check | 0.100 | 09:35:05 |
| Air Blank | 0.000 | 09:35:35 |
| Subject Test | 0.000 | 09:36:05 |
| Air Blank | 0.000 | 09:36:32 |
| Five Minute Wait | ABT* | |
| Air Blank | 0.000 | 09:37:13 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001108
10/26/2007

09:37:17

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 09:38:09 |
| Diagnostic Test | Pass | 09:38:36 |
| Air Blank | 0.000 | 09:39:04 |
| 0.100 Cal Check | RFI* | 09:39:16 |
| Air Blank | IPS** | 09:39:27 |

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

CM

3078
B4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001108
10/26/2007

09:39:39

CU
4088
TS4

Standard Lot#: 719801I
Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 09:40:36 |
| Diagnostic Test | Fail* | 09:40:39 |
| Air Blank | 0.000 | 09:41:08 |

* Analytical Stability Test

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number : 80-001108

10/26/2007

09:41:13

u

5078
By

Standard Lot#: 719801I

Last Changed By: C. CONLEY

#34517

QAS: T. GALLEGOS

#28847

TUCSON PD CRIME LAB

Operator: T. GALLEGOS

#28847

TUCSON PD CRIME LAB

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 09:42:00 |
| Diagnostic Test | Pass | 09:42:28 |
| Air Blank | 0.000 | 09:42:55 |
| 0.100 Cal Check | 0.100 | 09:43:14 |
| Air Blank | 0.000 | 09:43:43 |
| Subject Test | 0.281* | 09:46:52 |
| Air Blank | 0.000 | 09:47:26 |
| Air Blank | 0.000 | 09:47:54 |
| Subject Test | XXX** | 09:48:18 |
| Air Blank | 0.000 | 09:48:54 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-001108

10/26/2007

09:49:11

CM

Standard Lot#: 719801I


Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847

TUCSON PD CRIME LAB

6078
By

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 09:49:38 |
| Cal Check | 0.098 | 09:49:57 |
| Air Blank | 0.000 | 09:50:27 |



Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

OK
11-23-07
7 of 8
TSY

QA Specialist Terry Gallegos 20887 AGENCY Tucson Police Department
DATE 10-26-2007 TIME 0931
Intoxilyzer Serial # 80-001108 Location ODM

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-001108

10/26/2007

09:31:40

CU

8078

Standard Lot#: 7198011

Last Changed By: C. CONLEY #34517

QAS: T. GALLEGOS #28847

TUCSON PD CRIME LAB

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | 0.000 | 09:32:08 |
| Cal Check | 0.099 | 09:32:26 |
| Air Blank | 0.000 | 09:32:56 |


Operator's Signature