



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001108 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

*TS4
7-26-2016*

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7/21/2016 Cal 1 0.098 AC Cal 2 0.097 AC

Between the dates of 6/23/2016 and 7/21/2016 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on _____.
 The instrument was returned to service on _____.

William L Loomis

William L. Loomis


7/21/2016

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 21st, day of July, 2016.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
WLL
Bh 7-26-2016

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 7/21/2016 TIME 08:43
INTOXILYZER SERIAL # 80-001108 LOCATION ODM

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 097 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Will Loomis

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-001108
Core Version: 8105.48
07/21/2016

08:45:01

2 of 7
WLL
T84

Standard Lot#: 12415100A4
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:45:55
Diagnostic Test	Pass	08:46:30
Air Blank	0.000	08:46:58
0.100 Cal Check	0.098	08:47:19
Air Blank	0.000	08:47:48
Subject Test	0.000	08:48:06
Air Blank	0.000	08:48:33
Wait	ABT*	
Air Blank	0.000	08:49:08

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-001108
Core Version: 8105.48
07/21/2016

08:49:14

Standard Lot#: 12415100A4
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, SQAP

DOB: 11/11/1987

Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:50:07
Diagnostic Test	Pass	08:50:42
Air Blank	0.000	08:51:10
0.100 Cal Check	RFI*	08:51:23
Air Blank	IPS**	08:51:31

*RFI Detect

**Improper Sample

3 of 7
0 WLL
TSL

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-001108
Core Version: 8105.48
07/21/2016

08:51:36

4 of 7
0 WLL
T86

Standard Lot#: 12415100A4
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, SQAP

DOB: 11/11/1987

Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:52:27
Diagnostic Test	Pass	08:53:02
Air Blank	0.000	08:53:29
0.100 Cal Check	0.098	08:53:50
Air Blank	0.000	08:54:20
Subject Test	0.097*	08:57:27
Air Blank	0.000	08:57:59
Air Blank	0.000	08:58:27
Subject Test	XXX**	08:58:51
Air Blank	0.000	08:59:24

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-001108
Core Version: 8105.48
07/21/2016 08:59:48

5 of 7
6 WLL
T84

Standard Lot#: 12415100A4
Last Changed By: W. LOOMIS #53545

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:00:18
0.100 Cal Check	0.097	09:00:39
Air Blank	0.000	09:01:09

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
TSL
7.26.2016
WLL

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 7/21/2016 TIME 08:43
Intoxilyzer Serial # 80-001108 Location ODM

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Willie Loomis

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-001108
Core Version: 8105.48
07/21/2016 08:43:28

7 of 7
WLL

Standard Lot#: 12415100A4
Last Changed By: W. LOOMIS #53545

T84

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:43:57
0.100 Cal Check	0.098	08:44:18
Air Blank	0.000	08:44:48

Preventative Maintenance
Performed -- Timer Reset