



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
8.4.2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001163 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 07/29/10 Cal 1 0.098 AC Cal 2 0.099 AC

Between the dates of 07/01/10 and 07/29/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Cheryl K. Straub

Cheryl K. Straub


08/02/10

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 2, day of August, 2010.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

187
CS

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

T54
8.4.2010

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department
DATE 07/29/10 TIME 13:47
INTOXILYZER SERIAL # 80-001163 LOCATION CRIME LAB

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 099 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Cheryl Straub

237
CS

TSG

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-001163
07/29/2010 13:49:20

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:50:14
Diagnostic Test	Pass	13:50:42
Air Blank	0.000	13:51:09
0.100 Cal Check	0.099	13:51:28
Air Blank	0.000	13:51:58
Subject Test	0.000	13:52:16
Air Blank	ABT*	13:52:24
Air Blank	0.000	13:52:52

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

377
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-001163
07/29/2010

13:52:58

TJ4

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	13:53:31
Air Blank	IPS**	13:53:39

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

487
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-001163
07/29/2010

13:53:45

TJG

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:54:34
Diagnostic Test	Pass	13:55:01
Air Blank	0.000	13:55:29
0.100 Cal Check	0.099	13:55:47
Air Blank	0.000	13:56:17
Subject Test	0.067*	13:59:26
Air Blank	0.000	13:59:57
Air Blank	0.000	14:00:25
Subject Test	XXX**	14:00:44
Air Blank	0.000	14:01:18

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

577
08

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-001163
07/29/2010

14:01:39

D4

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	14:02:06
Cal Check	0.099	14:02:25
Air Blank	0.000	14:02:55



Operator's Signature

677
08

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

TS4
8.4.2010

QA Specialist Cheryl Straub AGENCY Tucson Police Department
DATE 07/29/10 TIME 13:47
Intoxilyzer Serial # 80-001163 Location CRIME Lab

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

777
CS

184

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-001163
07/29/2010 13:47:18

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	13:47:45
0.100 Cal Check	0.098	13:48:04
Air Blank	0.000	13:48:34
End of test		



Operator's Signature