



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
1306 W. Miracle Mile, Tucson, AZ 85705

SC  
5/19/14

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-001163 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5.1.2014 Cal 1 0.096 AC Cal 2 0.097 AC

Between the dates of 4.2.2014 and 5.1.2014 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.  
The instrument was returned to service on —.

Terry Gallegos

Terry S. Gallegos


5.15.2014

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 15<sup>th</sup>, day of May, 2014.

Elma Escarcega

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

10F7  
TJ4  
SC  
5/19/14

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 5.1.2014 TIME 1119

INTOXILYZER SERIAL # 80-001163 LOCATION ODM Sgt's Office

(  ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.
- (  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.
- (  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (  ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 097 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE 

2 of 7  
TS4  
SC

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-001163  
Core Version: 8105.48  
05/01/2014 11:21:07

Standard Lot#: 25913100A2  
Last Changed By: S. CLONE #50616

OAS: T. GALLEGOS #28847  
TUCSON PD

Operator: T. GALLEGOS #28847  
TUCSON PD

Subject: TEST, SOAP  
DOB: 11/11/1987  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:21:59
Diagnostic Test	Pass	11:22:37
Air Blank	0.000	11:23:04
0.100 Cal Check	0.098	11:23:25
Air Blank	0.000	11:23:54
Subject Test	0.000	11:24:11
Air Blank	0.000	11:24:39
Wait	ABT*	
Air Blank	0.000	11:25:26

\*Sequence Aborted

Not a Successfully  
Completed  
Test Sequence

3 of 7  
TS4  
SC

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-001163  
Core Version: 8105.48  
05/01/2014 11:25:45

Standard Lot#: 25913100A2  
Last Changed By: S. CIONE #50616

GAS: T. GALLEGOS #28847  
TUCSON PD

Operator: T. GALLEGOS #28847  
TUCSON PD

Subject: TEST, SOAP  
DOB: 11/11/1987  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:26:37
Diagnostic Test	Pass	11:27:14
Air Blank	RFI*	11:27:24
Air Blank	IPS**	11:27:36

\*RFI Detect  
\*\*Improper Sample

Not a Successfully  
Completed  
Test Sequence



4077  
B4  
5C

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-001163  
Core Version: 8105.48  
05/01/2014 11:28:18

Standard Lot#: 25913100A2  
Last Changed By: S. CIONE #50616

QAS: T. GALLEGOS #28847  
TUCSON PD

Operator: T. GALLEGOS #28847  
TUCSON PD

Subject: TEST, SOAP  
DOB: 11/11/1987  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:29:07
Diagnostic Test	Pass	11:29:45
Air Blank	0.000	11:30:12
0.100 Cal Check	0.097	11:30:33
Air Blank	0.000	11:31:02
Subject Test	0.136*	11:34:09
Air Blank	0.000	11:34:42
Air Blank	0.000	11:35:09
Subject Test	XXX**	11:35:34
Air Blank	0.000	11:36:08

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

5057  
TS4

5C

INTOX YZER 8000  
Location: TUCSON PD  
Serial Number: 80-001163  
Core Version: 8105.48  
05/01/2014 11:36:56

Standard Lot#: 25913100A2  
Last Changed By: S. CIONE #50616

OAS: T. GALLEGOS #28847  
TUCSON PD

Operator: T. GALLEGOS #28847  
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:37:28
0.100 Cal Check	0.097	11:37:49
Air Blank	0.000	11:38:18

Preventative Maintenance  
Performed -- Timer Reset

1

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

60f7  
TS4

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

SC  
5/19/14

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department  
DATE 5.1.2014 TIME 1119  
Intoxilyzer Serial # 80-001163 Location ODM Sgt's office

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 096 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

DPS Form Exh G-2 (Rev. 05-01)

7077  
TS4

SC

INTOXILYZER 8000  
Instrument Initialization  
10:17:35 05/01/2014

INTOXILYZER 8000  
Location: TUCSON PD  
Serial Number: 80-001163  
Core Version: 8105.48  
05/01/2014 11:19:16

-----  
Standard Lot#: 25913100A2  
Last Changed By: S. CIONE #50616  
-----

DAS: T. GALLEGOS #28847  
TUCSON PD  
-----

Operator: T. GALLEGOS #28847  
TUCSON PD CRIME LAB  
-----

Test	g/210L	Time
Air Blank	0.000	11:19:43
0.100 Cal Check	0.096	11:20:04
Air Blank	0.000	11:20:34

|