CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TSH
10-29-2014

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001163 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10/17/14 Cal 1 - AC Cal 2 - AC

Between the dates of - and - the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 10/17/14.
The instrument was returned to service on _____.

Sarah Cione

Sarah L. Cione

10/20/14

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 20th, day of October, 2014.

Elma Escarcega

Notary Public

OFFICIAL SEAL
ELMA ESCARCEGA
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Nov. 3, 2015

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 2 SC

TSH 10-29-2014

INTOXILYZER SERVICE RECORD

Date: 10/17/14

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-001163

Reason For Service: Initial Diagnostics during Warm-up - multiple failures. Won't go to "Ready Mode". Also unable to download records.

Date Out of Service: 10/17/14

Calibration Check completed prior to removal from service

Date: -

SQAP completed prior to removal from service

Date: -

Corrective Action: ☑ Sent to CMI, Inc.

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 2 sc

TSH

CMI INC.

Service Evaluation Form

This form MUST be completed and enclosed with instrument to be serviced.
Failure to complete and return this form may cause delays in service.

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number 105075 (contact Customer Service)

Name Terry Gallegos Phone: (520) 837-7677

Fax: (520) 791-4642 Email: Terry.Gallegos@tucsonaz.gov

2. Bill to Address:

Finance / Accounts Payable
PO Box 27450
Tucson, AZ 85726-7450

Ship to Address:

Tucson Police Department Crime Lab
1306 W. Miracle Mile
Tucson, AZ 85705

3. Serial Number: 80-001163 Instrument Model: I8000

4. Detailed Description of Problem:

Initial Diagnostics - Multiple failures (EEPROM + Analytical)
Can't download records

Hazardous Material Warning! – DO NOT return gas cylinders with instrument!

☐ I Authorize Repairs Up To: ☐ All ☐ $250 ☐ $500 ☐ $750 ☐ Other $_____

Purchase Order Number (attach a copy of P.O. if applicable) 34770

Authorized By:

Terry Gallegos Toxicology Coordinator
Name (Please Print) Title

Terry Gallegos 10·17·2014
Signature Date

Ship to:

CMI, Inc.
Attn: Service Dept.
316 East Ninth Street
Owensboro, KY 42303

☑ No, please send estimate before repairs are made.
Note: An estimate will be faxed before performing any repairs and may cause delays in service.
An evaluation fee ($79.00 or actual costs) will apply to estimates that are not repaired.