CERTIFICATION by CUSTODIAN of RECORDS

*Tucson Police Department Crime Laboratory*
*1306 W. Miracle Mile, Tucson, AZ 85705*

SC
2/16/15

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001163 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1.29.2015 Cal 1 0.097 AC Cal 2 0.096 AC

Between the dates of 12.31.2014 and 1.29.2015 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Terry Gallegos

Terry S. Gallegos

2.12.2015

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 12th, day of February, 2015.

Elma Escarcega

Notary Public

OFFICIAL SEAL
ELMA ESCARCEGA
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Nov. 3, 2015

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

1 of 7
TDH

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

SC
2/16/15

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Galleyos 28847 AGENCY Tucson Police Department

DATE 1-29-2015 TIME 1421

INTOXILYZER SERIAL # 80-001163 LOCATION DUI-HQ

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

( ✓ ) 1. Clock time check.
( ✓ ) 2. Date check.

OPERATIONAL TESTS

( ✓ ) 1. Alcohol-free subject test result 0.000 AC.
( ✓ ) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
( ✓ ) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
( ✓ ) 4. Standard Calibration Check standard 0.100 AC
Result: 0.096 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Terry Galleyos

DPS Form Exh G-4 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department*

2 of 7
TG

SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
01/29/2015 14:24:03

---

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

---

QAS: T. GALLEGOS #28847
TUCSON PD

---

Operator: T. GALLEGOS #28847
TUCSON PD

---

Subject: TEST,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 14:25:26 |
| Diagnostic Test | Pass | 14:26:01 |
| Air Blank | 0.000 | 14:26:28 |
| 0.100 Cal Check | 0.097 | 14:26:49 |
| Air Blank | 0.000 | 14:27:18 |
| Subject Test | 0.000 | 14:27:40 |
| Air Blank | 0.000 | 14:28:07 |
| Wait | ABT* | |
| Air Blank | 0.000 | 14:28:43 |

*Sequence Aborted

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7
TSG

SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
01/29/2015 14:29:54

---

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

---

QAS: T. GALLEGOS #28847
TUCSON PD

---

Operator: T. GALLEGOS #28847
TUCSON PD

---

Subject: TEST,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 14:30:44 |
| Diagnostic Test | Pass | 14:31:19 |
| Air Blank | 0.000 | 14:31:46 |
| 0.100 Cal Check | RFI* | 14:32:01 |
| Air Blank | IPS** | 14:32:10 |

*RFI Detect
**Improper Sample

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7
154
5c

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
01/29/2015 14:35:39

---

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

---

QAS: T. GALLEGOS #28847
TUCSON PD

---

Operator: T. GALLEGOS #28847
TUCSON PD

---

Subject: TEST,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 14:36:59 |
| Diagnostic Test | Pass | 14:37:34 |
| Air Blank | 0.000 | 14:38:01 |
| 0.100 Cal Check | 0.097 | 14:38:22 |
| Air Blank | 0.000 | 14:38:51 |
| Subject Test | 0.163* | 14:41:59 |
| Air Blank | 0.000 | 14:42:31 |
| Air Blank | 0.000 | 14:42:59 |
| Subject Test | XXX** | 14:43:33 |
| Air Blank | 0.000 | 14:44:07 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7
TSH

5C

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
01/29/2015 14:44:47

---

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

---

QAS: T. GALLEGOS #28847
TUCSON PD

---

Operator: T. GALLEGOS #28847
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 14:45:16 |
| 0.100 Cal Check | 0.096 | 14:45:37 |
| Air Blank | 0.000 | 14:46:07 |

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6 of 7
TSH

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

SC
2/16/15

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department

DATE 1.29.2015 TIME 1421

Intoxilyzer Serial # 80-001163 Location DUI-HQ

( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

( ✓ ) 4. Air blank completed.

( ✓ ) 5. Calibration check completed. Test results 0.097 AC.

( ✓ ) 6. Air blank completed.

( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Terry Gallegos

DPS Form Exh G-2 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

7 of 7
TSG

SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
01/29/2015 14:21:42

---

Standard Lot#: 11514100A2
Last Changed By: T. GALLEGOS #28847

---

QAS: T. GALLEGOS #28847
TUCSON PD

---

Operator: T. GALLEGOS #28847
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 14:22:28 |
| 0.100 Cal Check | 0.097 | 14:22:48 |
| Air Blank | 0.000 | 14:23:18 |

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.