CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001163 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TSH 7.27.2016

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 7-22-16 Cal 1 0.095 AC Cal 2 0.096 AC

Between the dates of 6-24-16 and 7-22-16 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Amy L. Autret

7-26-16

ADPS QAS Permit # 33718

SUBSCRIBED AND SWORN to before me this 26th, day of July, 2016.

Elma Escarcega

Notary Public

OFFICIAL SEAL
ELMA ESCARCEGA
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Nov. 3, 2019

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
JLA
TSH 7-27-2016

QA SPECIALIST Amy Autret 53546 AGENCY Tucson Police Department

DATE 7-22-16 TIME 1128

INTOXILYZER SERIAL # 80-001163 LOCATION crime laboratory

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
( ) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 095 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS _____

SIGNATURE _____

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
07/22/2016 11:30:44

Standard Lot#: 15615100A1
Last Changed By: W. LOOMIS #53545

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:31:36
Diagnostic Test	Pass	11:32:11
Air Blank	0.000	11:32:39
0.100 Cal Check	0.095	11:32:59
Air Blank	0.000	11:33:29
Subject Test	0.000	11:33:42
Air Blank	0.000	11:34:10
Wait	ABT*	
Air Blank	0.000	11:34:44

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
07/22/2016 11:34:48

Standard Lot#: 15615100A1
Last Changed By: W. LOOMIS #53545

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:35:36
Diagnostic Test	Pass	11:36:11
Air Blank	0.000	11:36:39
0.100 Cal Check	RFI*	11:36:53
Air Blank	IPS**	11:37:03

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
07/22/2016 11:37:10

Standard Lot#: 15615100A1
Last Changed By: W. LOOMIS #53545

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1987
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:38:00
Diagnostic Test	Pass	11:38:36
Air Blank	0.000	11:39:03
0.100 Cal Check	0.096	11:39:24
Air Blank	0.000	11:39:53
Subject Test	0.484*	11:43:01
Air Blank	0.000	11:43:34
Air Blank	0.000	11:44:01
Subject Test	XXX**	11:44:28
Air Blank	0.000	11:45:01

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
07/22/2016 11:45:14

Standard Lot#: 15615100A1
Last Changed By: W. LOOMIS #53545

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:45:42
0.100 Cal Check	0.095	11:46:03
Air Blank	0.000	11:46:33

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
TSG 7-27-2016

QA Specialist Amy Autret 53546 AGENCY Tucson Police Department

DATE 7-22-16 TIME 1128

Intoxilyzer Serial # 80-001163 Location Crime Laboratory

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.096 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature]

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001163
Core Version: 8105.48
07/22/2016 11:28:57

Standard Lot#: 15615100A1
Last Changed By: W. LOOMIS #53545

QAS: A. AUTRET #53546
TUCSON PD

Operator: A. AUTRET #53546
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:29:44
0.100 Cal Check	0.096	11:30:05
Air Blank	0.000	11:30:34

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.