



CERTIFICATION by CUSTODIAN of RECORDS  
Tucson Police Department Crime Laboratory

154

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-001165 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 2/15/07 Cal 1 0.099 AC Cal 2 0.099 AC

Between the dates of 1/19/07 and 2/15/07 the instrument was operating properly and accurately.


The instrument was taken out of service for repair on N/A.  
The instrument was returned to service on N/A.

Jennifer L. Coombs  
Jennifer L. Coombs

2/16/07

ADPS QAS Permit # 33120

SUBSCRIBED AND SWORN to before me this 16, day of February, 2007.



Kathy A. Kinney  
Notary Public

\*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

1 of 8  
jlc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

134

QA SPECIALIST Jennifer Coombs AGENCY Tucson Police Department  
DATE February 15, 2007 TIME 1445  
INTOXILYZER SERIAL # 80-001165 LOCATION OAS

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ( ✓ ) 1. Clock time check.
- ( ✓ ) 2. Date check.

OPERATIONAL TESTS

- ( ✓ ) 1. Alcohol-free subject test result 0. 000 AC.
- ( ✓ ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- ( ✓ ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed ✓✓  
Deficient sample printed.
- ( ✓ ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 099 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Jennifer L. Coombs

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-001165  
02/15/2007

14:47:23

BS4

2 of 8  
JL

---

Standard Lot#: 630601I  
Last Changed By: J. COOMBS #46160

---

QAS: J. COOMBS #46160  
TUCSON PD CRIME LAB

---

Operator: J. COOMBS #46160  
TUCSON PD CRIME LAB

---

Subject: SQAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	14:48:56
Diagnostic Test	Pass	14:49:23
Air Blank	0.000	14:49:50
0.100 Cal Check	0.100	14:50:09
Air Blank	0.000	14:50:39
Subject Test	0.000	14:50:54
Air Blank	0.000	14:51:21
Five Minute Wait	ABT*	
Air Blank	0.000	14:52:05

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-001165  
02/15/2007

14:52:14

BM

3 of 8  
file

---

Standard Lot#: 630601I  
Last Changed By: J. COOMBS #46160

---

QAS: J. COOMBS #46160  
TUCSON PD CRIME LAB

---

Operator: J. COOMBS #46160  
TUCSON PD CRIME LAB

---

Subject: SQAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	14:52:52
Air Blank	0.000	14:53:21

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-001165  
02/15/2007

14:53:26

Bh

4 of 8  
jlc

---

Standard Lot#: 630601I  
Last Changed By: J. COOMBS #46160

---

QAS: J. COOMBS #46160  
TUCSON PD CRIME LAB

---

Operator: J. COOMBS #46160  
TUCSON PD CRIME LAB

---

Subject: SQAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	14:54:14
Diagnostic Test	Pass	14:54:41
Air Blank	0.000	14:55:08
0.100 Cal Check	RFI*	14:55:26
Air Blank	0.000	14:55:56

\*RFI Detect

**Not a Successfully Completed Test Sequence**

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

5 of 8  
file

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-001165  
02/15/2007

14:56:02

Bh

Standard Lot#: 630601I  
Last Changed By: J. COOMBS #46160

QAS: J. COOMBS #46160  
TUCSON PD CRIME LAB

Operator: J. COOMBS #46160  
TUCSON PD CRIME LAB

Subject: SQAP, TEST  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:57:10
Diagnostic Test	Pass	14:57:38
Air Blank	0.000	14:58:05
0.100 Cal Check	0.099	14:58:24
Air Blank	0.000	14:58:54
Subject Test	0.214*	15:02:02
Air Blank	0.000	15:02:34
Air Blank	0.000	15:03:02
Subject Test	XXX**	15:03:16
Air Blank	0.000	15:03:49

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-001165  
02/15/2007 15:04:43

TS4

leaf 8  
JL

-----  
Standard Lot#: 630601I  
Last Changed By: J. COOMBS #46160  
-----

QAS: J. COOMBS #46160  
TUCSON PD CRIME LAB  
-----

Test	g/210L	Time
Air Blank	0.000	15:05:11
Cal Check	0.099	15:05:29
Air Blank	0.000	15:05:59

  
-----  
Operator's Signature

EXHIBIT G-2

7 of 8  
file

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

64

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Jennifer Coombs 46160 AGENCY Tucson Police Department  
DATE February 15, 2007 TIME 1445  
Intoxilyzer Serial # 80-001165 Location ODS

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 0999 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Jennifer L. Coombs

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-001165

02/15/2007

14:45:10

808  
file  
154

-----  
Standard Lot#: 630601I

Last Changed By: J. COOMBS #46160  
-----

QAS: J. COOMBS #46160

TUCSON PD CRIME LAB  
-----

Test	g/210L	Time
Air Blank	0.000	14:45:37
Cal Check	0.099	14:45:56
Air Blank	0.000	14:46:26

  
-----  
Operator's Signature