



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001199 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 5.24.2007 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 4.26.2007 and 5.24.2007 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on —.
The instrument was returned to service on —.


Terry S. Gallegos

Terry S. Gallegos

6.5.2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 5, day of June, 2007.



Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-1C-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

1 of 7
TSg
mm
6/9/07

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallegos 20847 AGENCY Tucson Police Department
DATE 5-24-2007 TIME 1533
INTOXILYZER SERIAL # 80-001199 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS


- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed ✓
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
05/24/2007

15:35:20

2 of 7
By ^{mm} 6-5-07

Standard Lot#: 627101G
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:36:10
Diagnostic Test	Pass	15:36:37
Air Blank	0.000	15:37:04
0.100 Cal Check	0.098	15:37:23
Air Blank	0.000	15:37:52
Subject Test	0.000	15:38:14
Air Blank	0.000	15:38:41
Five Minute Wait	ABT*	
Air Blank	0.000	15:39:42

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
05/24/2007 15:39:51

3 of 7
DJ
MM
6-5-07

Standard Lot#: 627101G
Last Changed By: T. GALLEGOS #28847
QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 100

Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:40:49
Diagnostic Test	Pass	15:41:17
Air Blank	0.000	15:41:44
Cal Check	0.098	15:42:03
Air Blank	0.000	15:42:32
Subject Test	RFI*	15:43:15
Air Blank	IPS**	15:43:24

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
05/24/2007 15:43:44

40f7
754^{Wm}
050?

Printed Label: 627101G
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Diagnostic Test	Pass	15:44:43
Air Blank	0.000	15:45:38
0.100 Cal Check	0.097	15:45:56
Air Blank	0.000	15:46:26
Subject Test	0.162*	15:49:34
Air Blank	0.000	15:50:06
Air Blank	0.000	15:50:33
Air Blank	0.000	15:51:00
Air Blank	0.000	15:51:33

*Deficient Sample
**Invalid Sample
Begin new deprivation period

No. 15:51:33 Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 87-101199

15:51:56

5077
TS4 ^{W4}₆₅₀₇

Standard Lot#: 627101G

Last Changed By: W. CALLEGOS #28847

QAS: W. CALLEGOS #28847

TUCSON PD CRIME LAB

	IL	Time
Air Blank	0.000	15:52:23
Cal Check	0.098	15:52:42
Air Blank	0.000	15:53:12


Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY


STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

6 of 7
TS4
6-5-07

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department
DATE 5.24.2007 TIME 1533
Intoxilyzer Serial # 80-001199 Location DUI HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C + 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
05/24/2007 15:33:30

7 of 7
TSG
2/23/07
6:50 PM

Standard Lot#: 627101G
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	15:33:57
Cal. Check	0.099	15:34:16
Air Blank	0.000	15:34:45



Operator's Signature