CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSH
7-1-09

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001199 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 6-25-09 Cal 1 0.100 AC Cal 2 0.098 AC

Between the dates of 6-4-09 and 6-25-09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

6-30-09

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 30th, day of June, 2009.

AMHernández

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 8

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 6.25.09 TIME 0844

INTOXILYZER SERIAL # 80-001199 LOCATION ODW DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.098 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
06/25/2009 08:45:41

2 of 8
CW
BY

Standard Lot#: 801002I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:46:26
Diagnostic Test	Pass	08:46:53
Air Blank	0.000	08:47:20
0.100 Cal Check	0.099	08:47:39
Air Blank	0.000	08:48:08
Subject Test	0.000	08:48:35
Air Blank	ABT*	08:48:42
Air Blank	0.000	08:49:10

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

3 of 8
Cu

B4

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
06/25/2009 08:50:36

Standard Lot#: 801002I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:51:12
Air Blank	IPS**	08:51:20

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

4 of 8
Ou

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
06/25/2009 08:58:33

Standard Lot#: 801002I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:00:05
Diagnostic Test	Pass	09:00:33
Air Blank	0.000	09:01:00
0.100 Cal Check	0.101	09:01:19
Air Blank	ABT*	09:01:38
Air Blank	0.000	09:02:07

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

5 of 8
CM

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-001199
06/25/2009 09:04:05

Standard Lot#: 801002I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:04:54
Diagnostic Test	Pass	09:05:21
Air Blank	0.000	09:05:49
0.100 Cal Check	0.099	09:06:07
Air Blank	0.000	09:06:37
Subject Test	0.135*	09:09:45
Air Blank	0.000	09:10:17
Air Blank	0.000	09:10:44
Subject Test	XXX**	09:11:15
Air Blank	0.000	09:11:48

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

6 of 8
Cu

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
06/25/2009 09:12:18

Standard Lot#: 801002I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:12:45
Cal Check	0.098	09:13:04
Air Blank	0.000	09:13:33

Clare Conley
Operator's Signature

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

7 of 8
Clm

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

BY
7-1-09

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 1-25-09 TIME 0844

Intoxilyzer Serial # 80-001199 Location ODW DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

8 of 8
CU

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
06/25/2009 08:44:16

Standard Lot#: 801002I
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

T54

Test	g/210L	Time
Air Blank	0.000	08:44:43
Cal Check	0.100	08:45:02
Air Blank	0.000	08:45:31

Clare Conley
Operator's Signature

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department