CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSL 1-6-2011

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-001199 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 1-6-11 Cal 1 0.100 AC Cal 2 0.100 AC

Between the dates of 12-9-10 and 1-6-11 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

1-6-11

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 6, day of January, 20 11.

Laura B Hodgson

Notary Public

OFFICIAL SEAL
LAURA B. HODGSON
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Aug. 26, 2013

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

1 of 7
CM
TSH 1.6.2011

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 1-6-11 TIME 0906

INTOXILYZER SERIAL # 80-001199 LOCATION ODW DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.100 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1827, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2807
CW
154

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
01/06/2011 09:08:35

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SOAP
DOB: 11/11/1111
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:09:25
Diagnostic Test	Pass	09:09:54
Air Blank	0.000	09:10:21
0.100 Cal Check	0.100	09:10:40
Air Blank	0.000	09:11:09
Subject Test	0.000	09:11:29
Air Blank	ABT*	09:11:35
Air Blank	0.000	09:12:04

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3:57
pm

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
01/06/2011 09:13:04

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SOAP
DOB: 11/11/1111
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	09:13:54
Air Blank	IPS**	09:14:02

*RFI Detect
**Improper Sample

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1377, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4507
154

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
01/06/2011 09:14:27

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SOAP
DOB: 11/11/1111
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:15:17
Diagnostic Test	Pass	09:15:46
Air Blank	0.000	09:16:13
0.100 Cal Check	0.100	09:16:32
Air Blank	0.000	09:17:01
Subject Test	0.100*	09:20:10
Air Blank	0.000	09:20:41
Air Blank	0.000	09:21:08
Subject Test	XXX**	09:21:26
Air Blank	0.000	09:21:59

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

Pursuant to section 28-1227, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7

154

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
01/06/2011 09:22:25

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

OAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:22:52
Cal Check	0.100	09:23:11
Air Blank	0.000	09:23:40

Clare Conley
Operator's Signature

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TSH 6.87 1.6.2011 CM

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 1-6-11 TIME 0906

Intoxilyzer Serial # 80-001199 Location ODW DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1827, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 & 7
Cm

TSG

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-001199
01/06/2011 09:06:44

Standard Lot#: 719919
Last Changed By: C. MORAND #24219

OAS: C. CONLEY #34517
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:07:11
Cal Check	0.100	09:07:30
Air Blank	0.000	09:07:59

Blare Conley
Operator's Signature

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.