



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
1306 W. Miracle Mile, Tucson, AZ 85705

TS4  
6-27-2012

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-001199 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 6/21/12 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 5/24/12 and 6/21/12 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.  
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione


6/21/12

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 21<sup>st</sup>, day of June, 2012.

Elma Escarcega

Notary Public



\*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

1 of 7 sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

BS  
6-27-2012

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 6/21/12 TIME 1012

INTOXILYZER SERIAL # 80-001199 LOCATION DUI HQ

(  ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.
- (  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.
- (  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed.  
Deficient sample printed.
- (  ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 098 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

---



---



---



---

SIGNATURE Sarah Cione

2 of 7<sub>sc</sub>

154

INTOXILYZER 8000  
 Location: TUCSON POLICE DEPT  
 Serial Number: 80-001199  
 Core Version: 8105.48  
 06/21/2012 10:15:05

---

Standard Lot#: 23511100A2  
 Last Changed By: J. HAWKINS #42401

---

QAS: S. CIONE #50616  
 TUCSON PD

---

Operator: S. CIONE #50616  
 TUCSON PD

---

Subject: SQAP, SQAP  
 DOB: 11/11/1980  
 Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 10:15:48 |
| Diagnostic Test | Pass   | 10:16:23 |
| Air Blank       | 0.000  | 10:16:51 |
| 0.100 Cal Check | 0.100  | 10:17:11 |
| Air Blank       | 0.000  | 10:17:41 |
| Subject Test    | 0.000  | 10:18:10 |
| Air Blank       | 0.000  | 10:18:38 |
| Wait            | ABT*   |          |
| Air Blank       | 0.000  | 10:19:48 |

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

3 of 7 sc  
TS4

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-001199  
Core Version: 8105.48  
06/21/2012 10:20:50

---

Standard Lot#: 23511100A2  
Last Changed By: J. HAWKINS #42401

---

QAS: S. CIONE #50616  
TUCSON PD

---

Operator: S. CIONE #50616  
TUCSON PD

---

Subject: SQAP, SQAP  
DOB: 11/11/1980  
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 10:21:35 |
| Diagnostic Test | Pass   | 10:22:10 |
| Air Blank       | 0.000  | 10:22:37 |
| 0.100 Cal Check | 0.099  | 10:22:58 |
| Air Blank       | 0.000  | 10:23:27 |
| Subject Test    | RFI*   | 10:23:37 |
| Air Blank       | IPS**  | 10:23:45 |

\*RFI Detect  
\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-001199  
Core Version: 8105.48  
06/21/2012 10:26:58

4 of 7sc

B4

Standard Lot#: 23511100A2  
Last Changed By: J. HAWKINS #42401

QAS: S. CIONE #50616  
TUCSON PD

Operator: S. CIONE #50616  
TUCSON PD

Subject: SQAP, SQAP  
DOB: 11/11/1980  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 10:27:40 |
| Diagnostic Test | Pass   | 10:28:15 |
| Air Blank       | 0.000  | 10:28:42 |
| 0.100 Cal Check | 0.099  | 10:29:03 |
| Air Blank       | 0.000  | 10:29:33 |
| Subject Test    | 0.047* | 10:32:40 |
| Air Blank       | 0.000  | 10:33:11 |
| Air Blank       | 0.000  | 10:33:39 |
| Subject Test    | XXX**  | 10:34:03 |
| Air Blank       | 0.000  | 10:34:37 |

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

5 of 7sc

TSL

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-001199  
Core Version: 8105.48  
06/21/2012 10:35:40

-----  
Standard Lot#: 23511100A2  
Last Changed By: J. HAWKINS #42401  
-----

QAS: S. CIONE #50616  
TUCSON PD  
-----

Operator: S. CIONE #50616  
TUCSON PD  
-----

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 10:36:09 |
| 0.100 Cal Check | 0.098  | 10:36:29 |
| Air Blank       | 0.000  | 10:36:59 |

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2

6 of 7 sc  
B4  
6-27-2012

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department  
DATE 6/21/12 TIME 1012  
Intoxilyzer Serial # 80-001199 Location DUI HQ

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 099 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TS4

INTOXILYZER 8000  
 Location: TUCSON POLICE DEPT  
 Serial Number: 80-001199  
 Core Version: 8105.48  
 06/21/2012 10:12:56

-----  
 Standard Lot#: 23511100A2  
 Last Changed By: J. HAWKINS #42401  
 -----

QAS: S. CIONE #50616  
 TUCSON PD  
 -----

Operator: S. CIONE #50616  
 TUCSON PD  
 -----

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 10:13:25 |
| 0.100 Cal Check | 0.099  | 10:13:46 |
| Air Blank       | 0.000  | 10:14:15 |

Preventative Maintenance  
 Performed -- Timer Reset



TUCSON POLICE DEPARTMENT  
CRIME LABORATORY

INTOXILYZER SERVICE RECORD  
FOR  
CALIBRATION GAS TANK CHANGE

Tdy  
6-27-2012

Date: 5/26/12  
QAS: J HAWKINS 42401  
Intoxilyzer Serial Number: 80-001199  
New Lot Number: 23511100A2  
New Expiration Date: 11/1/2013

- (✓) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 2. Remove the old gas tank and install the new one.
- (✓) 3. Hit the "ESC" key twice and swipe the QAS card.
- (✓) 4. Go to the "M-Maintenance" menu and select "S-Standard Change."  
Select "D" for dry gas and enter the following information:  
Enter Target Value: 0.100  
Enter Control Lot #: From the tank label  
Enter Exp Date: From the tank label (mm/dd/yyyy)  
Edit info Y/N? Y or N  
Update Std Info? Y ENTER  
"Please wait saving..."
- (✓) 5. Hit the "ESC" key twice.
- (✓) 6. Go to the "M-Maintenance" menu and select "T-Control Testing."  
Select "D" for dry control test. After selection is made press ENTER.
- (✓) 7. Air blank completed.
- (✓) 8. Calibration check completed. Test results 0.101 / 100 AC.
- (✓) 9. Air blank completed.
- (✓) 10. Remove and sign printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 42401

TS4

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-001199  
Core Version: 8105.48  
05/26/2012 19:25:15

Standard Lot#: 23511100A2  
Last Changed By: J. HAWKINS #42401

OAS: S. CIONE #50616  
TUCSON PD

Operator: J. HAWKINS #42401  
TUCSON PD

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 19:25:46 |
| 0.100 Cal Check | 0.100  | 19:26:06 |
| Air Blank       | 0.000  | 19:26:36 |

Preventative Maintenance NOT  
Performed -- Timer NOT Reset

*J. Hawkins #42401*

INTOXILYZER 6000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-001199  
Core Version: 8105.48  
05/26/2012 19:22:22

Standard Lot#: 23511100A2  
Last Changed By: J. HAWKINS #42401

OAS: S. CIONE #50616  
TUCSON PD

Operator: J. HAWKINS #42401  
TUCSON PD

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 19:22:53 |
| 0.100 Cal Check | 0.101  | 19:23:14 |
| Air Blank       | 0.000  | 19:23:43 |

Preventative Maintenance NOT  
Performed -- Timer NOT Reset

*J. Hawkins #42401*