



CERTIFICATION by CUSTODIAN of RECORDS  
Tucson Police Department Crime Laboratory

OK  
11-23-07

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-002201 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10-19-2007 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 9-19-2007 and 10-19-2007 the instrument was operating properly and accurately.


The instrument was taken out of service for repair on —.  
The instrument was returned to service on —.

Terry S. Gallegos

11-21-2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 21, day of November, 2007.



Notary Public

\*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

CU  
11-23-07

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

10 of 7  
TS4

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Terry Gallegos 28947 AGENCY Tucson Police Department

DATE 10-19-2007 TIME 1527

INTOXILYZER SERIAL # 80-002201 LOCATION DUI HQ (room 12)

(  ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.
- (  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.
- (  ) 2. Error recognition logic system functioning. ✓✓✓  
Not a Successfully Completed Test Sequence printed
- (  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (  ) 4. Standard Calibration Check standard 0. 100 AC  
Result: 0. 099 AC

Instrument is operating properly and accurately. YES  NO

COMMENTS

---



---



---



---

SIGNATURE

Terry Gallegos

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-002201  
10/19/2007 15:28:50

Ch  
2 of 7  
154

Standard Lot#: 706602I  
Last Changed By: J. OLIVARES #36552

QAS: T. GALLEGOS #28847  
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847  
TUCSON PD CRIME LAB

Subject: SQAP, SQAP  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:29:44
Diagnostic Test	Pass	15:30:12
Air Blank	0.000	15:30:39
0.100 Cal Check	0.100	15:30:58
Air Blank	0.000	15:31:27
Subject Test	0.000	15:31:44
Air Blank	0.000	15:32:12
Five Minute Wait	ABT*	
Air Blank	0.000	15:32:47

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-002201  
10/19/2007 15:32:58

*cu*  
*30 of 7*  
*TS4*

Standard Lot#: 706602I  
Last Changed By: J. OLIVARES #36552

QAS: T. GALLEGOS #28847  
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847  
TUCSON PD CRIME LAB

Subject: SQAP, SQAP  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:33:48
Diagnostic Test	Pass	15:34:16
Air Blank	RFI*	15:34:38
Air Blank	IPS**	15:34:49

\*RFI Detect  
\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-002201  
10/19/2007 15:35:16

Cu  
4 of 7  
Ph

Standard Lot#: 706602I  
Last Changed By: J. OLIVARES #36552

QAS: T. GALLEGOS #28847  
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847  
TUCSON PD CRIME LAB

Subject: SQAP, SQAP  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:36:08
Diagnostic Test	Pass	15:36:35
Air Blank	0.000	15:37:02
0.100 Cal Check	0.100	15:37:21
Air Blank	0.000	15:37:51
Subject Test	0.068*	15:40:59
Air Blank	0.000	15:41:31
Air Blank	0.000	15:41:58
Subject Test	XXX**	15:42:17
Air Blank	0.000	15:42:51

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-002201  
10/19/2007 15:43:41

cu  
5077  
TS4

-----  
Standard Lot#: 706602I  
Last Changed By: J. OLIVARES #36552  
-----

QAS: T. GALLEGOS #28847  
TUCSON PD CRIME LAB  
-----

Test	g/210L	Time
Air Blank	0.000	15:44:08
Cal Check	0.099	15:44:27
Air Blank	0.000	15:44:57


  
-----  
Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY


STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

CM  
11-23-07  
6 of 7  
TS4

QA Specialist Terry Gallegos 28847 AGENCY Tucson Police Department  
DATE 10-19-2007 TIME 1527  
Intoxilyzer Serial # 80-002201 Location DUI HQ (Code 12)

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 100 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-002201

10/19/2007

15:27:26

cu  
7 of 7  
TSh


-----  
Standard Lot#: 706602I

Last Changed By: J. OLIVARES #36552

-----  
QAS: T. GALLEGOS #28847

TUCSON PD CRIME LAB

-----  
Test                      g/210L                      Time  
-----  
Air Blank                  0.000                      15:27:54  
Cal Check                  0.100                      15:28:13  
Air Blank                  0.000                      15:28:42

  
-----  
Operator's Signature