CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSY 10-2-09

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002201 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 09/17/09 Cal 1 0.097 AC Cal 2 0.096 AC

Between the dates of 08/21/09 and 09/17/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Cheryl K. Straub

09/29/09

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 29th, day of September, 2009.

AMHernández

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

177 a

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TSY 10-2-09

QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police Department

DATE 09/17/09 TIME 07:54

INTOXILYZER SERIAL # 80-002201 LOCATION CRIME LAB

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 096 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS INVALID SAMPLE TEST #1: I RAN OUT OF BREATH WHILE PERFORMING THE TEST AND IT WAS ACCEPTED AS A SAMPLE - THE INVALID TEST WAS REPEATED

SIGNATURE Cheryl Straub

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

217
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002201
09/17/2009 07:56:10

154

Standard Lot#: 826001I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:57:35
Diagnostic Test	Pass	07:58:02
Air Blank	0.000	07:58:29
0.100 Cal Check	0.097	07:58:48
Air Blank	0.000	07:59:18
Subject Test	0.000	07:59:52
Air Blank	ABT*	07:59:59
Air Blank	0.000	08:00:27

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

377
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002201
09/17/2009 08:00:36

Standard Lot#: 826001I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:01:33
Diagnostic Test	Pass	08:02:00
Air Blank	RFI*	08:02:09
Air Blank	IPS**	08:02:17

*RFI Detect
**Improper Sample

134

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

477
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002201
09/17/2009 08:02:50

154

Standard Lot#: 826001I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:03:43
Diagnostic Test	Pass	08:04:10
Air Blank	0.000	08:04:38
0.100 Cal Check	0.097	08:04:57
Air Blank	0.000	08:05:26
Subject Test	0.207*	08:08:35
Air Blank	0.000	08:09:08
Air Blank	0.000	08:09:35
Subject Test	0.186	08:10:07
Air Blank	0.000	08:10:42
Five Minute Wait		
Air Blank	0.000	08:15:34
Subject Test	XXX**	08:16:07
Air Blank	0.000	08:16:41

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

527
cs

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
09/17/2009 08:17:17

Standard Lot#: 826001I
Last Changed By: S. CIONE #50616

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:17:44
Cal Check	0.096	08:18:03
Air Blank	0.000	08:18:33

154

Operator's Signature

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

677
CS

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

154
10-2-09

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department

DATE 09/17/09 TIME 07:54

Intoxilyzer Serial # 80-002201 Location CRIME LAB

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

777 CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002201
09/17/2009 07:54:01

Standard Lot#: 826001I
Last Changed By: S. CIONE #50616

154

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	07:54:28
Cal Check	0.097	07:54:47
Air Blank	0.000	07:55:17

Operator's Signature

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.