CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TSH
4.29.2014

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002201 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4/25/14 Cal 1 0.100 AC Cal 2 0.100 AC

Between the dates of 3/28/14 and 4/25/14 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione

4/28/14

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 28th, day of April, 2014.

Elma Escarcega

Notary Public

OFFICIAL SEAL
ELMA ESCARCEGA
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp Nov. 3, 2015

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7 sc
TSH
4.29.2014

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 4/25/14 TIME 1032

INTOXILYZER SERIAL # 80-002201 LOCATION Crime Lab

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.100 AC

Instrument is operating properly and accurately. YES ✓ NO____

COMMENTS

SIGNATURE Sarah Cione

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 7
SC
BH

INTOXILYZER 8000
Location: TUCSON PD U OF A
Serial Number: 80-002201
Core Version: 8105.48
04/25/2014 10:33:41

Standard Lot#: 30412100A1
Last Changed By: R. SZELEWSKI #46927

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1980
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:34:26
Diagnostic Test	Pass	10:35:01
Air Blank	0.000	10:35:29
0.100 Cal Check	0.100	10:35:50
Air Blank	0.000	10:36:19
Subject Test	0.000	10:36:44
Air Blank	0.000	10:37:12
Wait	ABT*	
Air Blank	0.000	10:37:51

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7 SC
BH

INTOXILYZER 8000
Location: TUCSON PD U OF A
Serial Number: 80-002201
Core Version: 8105.48
04/25/2014 10:38:09

Standard Lot#: 30412100A1
Last Changed By: R. SZELEWSKI #46927

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1980
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:38:50
Diagnostic Test	Pass	10:39:25
Air Blank	0.000	10:39:53
0.100 Cal Check	0.101	10:40:13
Air Blank	0.000	10:40:43
Subject Test	RFI*	10:40:54
Air Blank	IPS**	10:41:01

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7 sc
BG

INTOXILYZER 8000
Location: TUCSON PD U OF A
Serial Number: 80-002201
Core Version: 8105.48
04/25/2014 10:41:11

Standard Lot#: 30412100A1
Last Changed By: R. SZELEWSKI #46927

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1980
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:42:13
Diagnostic Test	Pass	10:42:48
Air Blank	0.000	10:43:16
0.100 Cal Check	0.100	10:43:36
Air Blank	0.000	10:44:06
Subject Test	0.197*	10:47:14
Air Blank	0.000	10:47:46
Air Blank	0.000	10:48:13
Subject Test	XXX**	10:48:36
Air Blank	0.000	10:49:09

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7 SC

B4

INTOXILYZER 8000
Location: TUCSON PD U OF A
Serial Number: 80-002201
Core Version: 8105.48
04/25/2014 10:49:44

Standard Lot#: 30412100A1
Last Changed By: R. SZELEWSKI #46927

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:50:13
0.100 Cal Check	0.100	10:50:34
Air Blank	0.000	10:51:03

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

6 of 7 Sc
BG
4.29.2014

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department

DATE 4/25/14 TIME 1032

Intoxilyzer Serial # 80-002201 Location Crime Lab

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

7 of 7 sc
BJ

INTOXILYZER 8000
Location: TUCSON PD U OF A
Serial Number: 80-002201
Core Version: 8105.48
04/25/2014 10:32:12

Standard Lot#: 30412100A1
Last Changed By: R. SZELEWSKI #46927

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:32:41
0.100 Cal Check	0.100	10:33:01
Air Blank	0.000	10:33:31

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.