



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

TS4
9-3-2014

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002201 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 8/14/14 Cal 1 0.100 AC Cal 2 0.099 AC

Between the dates of 7/17/14 and 8/14/14 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione

8/19/14


ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 19th day of August, 2014.

Elma Escarcega

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5



Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7
sc
PJ
9-3-2014

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 8/14/14 TIME 1044

INTOXILYZER SERIAL # 80-002201 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 099 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002201
Core Version: 8105.48
08/14/2014

10:47:20

2 of 7_{sc}

TS4

Standard Lot#: 25913100A2
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:48:41 |
| Diagnostic Test | Pass | 10:49:16 |
| Air Blank | 0.000 | 10:49:43 |
| 0.100 Cal Check | 0.100 | 10:50:04 |
| Air Blank | 0.000 | 10:50:34 |
| Subject Test | 0.000 | 10:51:08 |
| Air Blank | 0.000 | 10:51:35 |
| Wait | ABT* | |
| Air Blank | 0.000 | 10:52:57 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002201
Core Version: 8105.48
08/14/2014 10:53:35

3 of 7
sc
TS4

Standard Lot#: 25913100A2
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:54:40 |
| Diagnostic Test | Pass | 10:55:15 |
| Air Blank | 0.000 | 10:55:43 |
| 0.100 Cal Check | 0.100 | 10:56:03 |
| Air Blank | 0.000 | 10:56:33 |
| Subject Test | RFI* | 10:57:09 |
| Air Blank | IPS** | 10:57:17 |

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002201
Core Version: 8105.48
08/14/2014

11:01:34

4 of 7 sc
B4

Standard Lot#: 25913100A2
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1980
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 11:02:27 |
| Diagnostic Test | Pass | 11:03:02 |
| Air Blank | 0.000 | 11:03:30 |
| 0.100 Cal Check | 0.100 | 11:03:50 |
| Air Blank | 0.000 | 11:04:20 |
| Subject Test | 0.260* | 11:07:27 |
| Air Blank | 0.000 | 11:07:59 |
| Air Blank | 0.000 | 11:08:27 |
| Subject Test | XXX** | 11:08:58 |
| Air Blank | 0.000 | 11:09:32 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002201
Core Version: 8105.48
08/14/2014 11:11:08

5 of 7
sc
Bq

Standard Lot#: 25913100A2
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 11:11:36 |
| 0.100 Cal Check | 0.099 | 11:11:57 |
| Air Blank | 0.000 | 11:12:26 |

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
5c
D4
9.3.2014

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 8/14/14 TIME 1044
Intoxilyzer Serial # 80-002201 Location DUT HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002201
Core Version: 8105.48
08/14/2014 10:44:59

7 of 7
sc
Bl

Standard Lot#: 25913100A2
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:45:28 |
| 0.100 Cal Check | 0.100 | 10:45:48 |
| Air Blank | 0.000 | 10:46:18 |

Preventative Maintenance
Performed -- Timer Reset



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 2
SC
BY
9-3-2014

INTOXILYZER SERVICE RECORD
FOR
CALIBRATION GAS TANK CHANGE

Date: 8-11-14
QAS: J. OLIVARES #36552
Intoxilyzer Serial Number: 80-002201
New Lot Number: 2591310042
10-01-15
New Expiration Date: _____

- (V) 1. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (V) 2. Remove the old gas tank and install the new one.
- (V) 3. Hit the "ESC" key twice and swipe the QAS card.
- (V) 4. Go to the "M-Maintenance" menu and select "S-Standard Change."
Select "D" for dry gas and enter the following information:
Enter Target Value: 0.100
Enter Control Lot #: From the tank label
Enter Exp Date: From the tank label (mm/dd/yyyy)
Edit info Y/N? Y or N
Update Std Info? Y ENTER
"Please wait saving..."
- (V) 5. Hit the "ESC" key twice.
- (V) 6. Go to the "M-Maintenance" menu and select "T-Control Testing."
Select "D" for dry control test. After selection is made press ENTER.
- (V) 7. Air blank completed. 0.000
- (V) 8. Calibration check completed. Test results 0.100 AC.
- (V) 9. Air blank completed. 0.000
- (V) 10. Remove and sign printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #36552

2 of 2 sc
TSY

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002201
Core Version: 8105.48
08/11/2014 21:53:59

Standard Lot#: 25913100A2
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Operator: J. OLIVARES #36552
TUCSON PD

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 21:54:35 |
| 0.100 Cal Check | 0.100 | 21:54:56 |
| Air Blank | 0.000 | 21:55:26 |

Preventative Maintenance NOT
Performed -- Timer NOT Reset

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002201
Core Version: 8105.48
08/11/2014 21:58:20

Standard Lot#: 25913100A2
Last Changed By: J. OLIVARES #36552

QAS: S. CIONE #50616
TUCSON PD

Operator: J. OLIVARES #36552
TUCSON PD

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 21:58:50 |
| 0.100 Cal Check | 0.100 | 21:59:11 |
| Air Blank | 0.000 | 21:59:41 |

Preventative Maintenance NOT
Performed -- Timer NOT Reset