



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002261 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

154
11-4-08

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10/23/08 Cal 1 0.103 AC Cal 2 0.101 AC

Between the dates of 9/25/08 and 10/23/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -
The instrument was returned to service on -

Sarah Cione

Sarah L. Cione


11/3/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 3rd day of November, 2008.

A.M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 8 sc

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TS4
11-4-08

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 10/23/08 TIME 8:35

INTOXILYZER SERIAL # 80-002261 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 101 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 8_{sc}

INTOXILYZER 8000
Location: CMI INC
Serial Number : 80-002261
10/23/2008

08:37:22

TS4

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:38:17
Diagnostic Test	Pass	08:38:45
Air Blank	0.000	08:39:12
0.100 Cal Check	0.102	08:39:31
Air Blank	0.000	08:40:00
Subject Test	0.000	08:41:05
Air Blank	0.000	08:41:32
Five Minute Wait	ABT*	
Air Blank	0.000	08:43:05

*Sequence Aborted

Not a Successfully Completed Test Sequence

3 of 8 sc

By

INTOXILYZER 8000
Location: CMI INC
Serial Number : 80-002261
10/23/2008

08:43:20

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:44:22
Air Blank	0.000	08:44:50

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CMI INC
Serial Number : 80-002261
10/23/2008

08:45:16

TS4

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	08:45:55
Air Blank	0.000	08:46:23

*Improper Sample

Not a Successfully Completed Test Sequence

5 of 8
sc

BM

INTOXILYZER 8000
Location: CMI INC
Serial Number : 80-002261
10/23/2008

08:47:44

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:48:41
Diagnostic Test	Pass	08:49:08
Air Blank	0.000	08:49:36
0.100 Cal Check	0.102	08:49:54
Air Blank	0.000	08:50:24
Subject Test	0.129*	08:53:32
Air Blank	0.000	08:54:05
Air Blank	0.000	08:54:32
Subject Test	XXX**	08:55:47
Air Blank	0.000	08:56:21

*Deficient Sample
 **Invalid Sample
 Begin new deprivation period

Not a Successfully Completed Test Sequence

6 of 8

INTOXILYZER 8000
Location: CMI INC
Serial Number: 80-002261
10/23/2008 08:57:15

BY

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:57:43
Cal Check	0.101	08:58:01
Air Blank	0.000	08:58:31

Sarah Cione
Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

7 of 8
SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TSK
11.4.08

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 10/23/08 TIME 8:35
Intoxilyzer Serial # 80-002261 Location DUI HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 103 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

8 of PSC

INTOXILYZER 8000
Location: CMI INC
Serial Number: 80-002261
10/23/2008 08:35:28

TS4

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:35:56
Cal Check	0.103	08:36:14
Air Blank	0.000	08:36:44

Sarah Cione
Operator's Signature