CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSL 9.21.2010

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002261 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 9/16/10 Cal 1 0.099 AC Cal 2 0.099 AC

Between the dates of 8/19/10 and 9/16/10 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione

9/20/10

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 20th, day of September, 2010.

A.M. Hernandez

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 8 SC

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TSH
9.21.2010

QA SPECIALIST Sarah Ciora AGENCY Tucson Police Department

DATE 9/16/10 TIME 1103

INTOXILYZER SERIAL # 80-002261 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.099 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS _____

SIGNATURE Sarah Ciora

DPS Form Exh G-4 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 8 SC

184

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
09/16/2010 11:05:48

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:06:50
Diagnostic Test	Pass	11:07:18
Air Blank	0.000	11:07:45
0.100 Cal Check	0.100	11:08:04
Air Blank	0.000	11:08:33
Subject Test	0.000	11:08:54
Air Blank	0.000	11:09:22
Five Minute Wait	ABT*	
Air Blank	0.000	11:09:58

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 8 SC

TSH

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
09/16/2010 11:13:28

Standard Lot#: 719919
Last Changed By: S. CIONE - #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	11:13:57
Air Blank	IPS**	11:14:05

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 8 SC

154

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
09/16/2010 11:31:05

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:31:51
Diagnostic Test	Pass	11:32:18
Air Blank	0.000	11:32:46
0.100 Cal Check	0.099	11:33:04
Air Blank	0.000	11:33:34
Subject Test	0.100*	11:36:42
Air Blank	0.000	11:37:14
Air Blank	0.000	11:37:41
Subject Test	ABT**	11:38:49
Air Blank	0.000	11:39:17

*Deficient Sample
**Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 8 SC

TSH

INTOXILYZER 8000
Location: TUCSON PD
Serial Number : 80-002261
09/16/2010 11:14:51

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:15:35
Diagnostic Test	Pass	11:16:03
Air Blank	0.000	11:16:30
0.100 Cal Check	0.100	11:16:49
Air Blank	0.000	11:17:18
Subject Test	XXX*	11:20:25
Air Blank	0.000	11:20:58

*Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

6 of 8 SC

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002261
09/16/2010 11:21:33

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:22:00
Cal Check	0.099	11:22:19
Air Blank	0.000	11:22:49

TS4

Sarah Cione
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

7 of 8 SC

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TSH 9.21.2010

QA Specialist Sarah Cione AGENCY Tucson Police Department

DATE 9/16/10 TIME 1103

Intoxilyzer Serial # 80-002261 Location DUI HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

8 of 8 SC

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002261
09/16/2010 11:03:48

Standard Lot#: 719919
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	11:04:15
Cal Check	0.099	11:04:34
Air Blank	0.000	11:05:03

Sarah Cione
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.