



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

Ch
11-23-07

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002262 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: N/A Cal 1 _____ AC Cal 2 _____ AC

Between the dates of 9.19.2007 and 10.17.2007 the instrument was operating properly and accurately.


The instrument was taken out of service for repair on 10.19.2007.
The instrument was returned to service on —.

Terry S. Gallegos

11.21.2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 21, day of November, 2007.



Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

CU
11-23-07

INTOXILYZER SERVICE RECORD

10f1
T86

Date: 10.19.2007

QAS: Terry Gallegos 28847

Intoxilyzer Serial Number: 80-002262

Reason For Service: Officer Stacy Lee (ODW) called on 10.17.2007 to say that the Intoxilyzer did not work at all. It was picked up and brought to the lab on 10.19.2007. The breath hose does not heat and the instrument will not go to "Ready" mode.

Date Out of Service: 10.19.2007

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action:

Sent to CMI, Inc.

COBRA download prior to shipping.