



CERTIFICATION by CUSTODIAN of RECORDS
 Tucson Police Department Crime Laboratory
 270 S. Stone Avenue Tucson, AZ 85701-1917

TS4
 12-22-08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002262 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 12/18/08 Cal 1 0.100 AC Cal 2 0.100 AC

Between the dates of 11/20/08 and 12/18/08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
 The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione


12/22/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 22, day of December, 2008.

A. M. Hernandez

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

1 of 75c

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

TSY
12-22-08

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 12/18/08 TIME P:58

INTOXILYZER SERIAL # 80-002262 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Sarah Cione

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

2 of 75c

B4

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002262
12/18/2008

09:00:22

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP,TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:01:13
Diagnostic Test	Pass	09:01:40
Air Blank	0.000	09:02:08
0.100 Cal Check	0.100	09:02:26
Air Blank	0.000	09:02:56
Subject Test	0.000	09:03:31
Air Blank	0.000	09:03:58
Five Minute Wait	ABT*	
Air Blank	0.000	09:04:35

*Sequence Aborted

Not a Successfully Completed Test Sequence

3 of 7sc

124

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002262
12/18/2008

09:05:43

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Operator: S. CIONE #50616
TUCSON PD

Subject: SQAP, TEST
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	09:06:20
Air Blank	IPS**	09:06:29

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7 SC

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002262
12/18/2008

09:07:19

TSY

Standard Lot#: 735401G	
Last Changed By: S. CIONE	#50616
QAS: S. CIONE	#50616
TUCSON PD	
Operator: S. CIONE	#50616
TUCSON PD	

Subject: SQAP, TEST
 DOB: 11/11/1111
 Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:08:06
Diagnostic Test	Pass	09:08:33
Air Blank	0.000	09:09:01
0.100 Cal Check	0.101	09:09:19
Air Blank	0.000	09:09:49
Subject Test	0.093*	09:12:57
Air Blank	0.000	09:13:29
Air Blank	0.000	09:13:56
Subject Test	XXX**	09:14:20
Air Blank	0.000	09:14:53

*Deficient Sample
 **Invalid Sample
 Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002262
12/18/2008

09:15:25

5 of 7 sc

124

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	09:15:53
Cal Check	0.100	09:16:11
Air Blank	0.000	09:16:41

Sarah Cione
Operator's Signature

6 of 7 sc

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

TSY
12-22-08

QA Specialist Sarah Cione AGENCY Tucson Police Department
DATE 12/18/08 TIME 8:58
Intoxilyzer Serial # 80-002262 Location DUTHQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 0.100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

7 of 7sc

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002262
12/18/2008 08:58:37

134

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: S. CIONE #50616
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	08:59:04
Cal Check	0.100	08:59:23
Air Blank	0.000	08:59:53

Sarah Cione
Operator's Signature