CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory
270 S. Stone Avenue Tucson, AZ 85701-1917

TSH 10-2-09

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002262 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 09/17/09 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 08/20/09 and 09/17/09 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Cheryl K. Straub

09/29/09

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 29th, day of September, 2009.

A.M. Hernández

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

197 CS

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

BY 10-2-09

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Cheryl Straub AGENCY Tucson Police Department

DATE 09/17/09 TIME 10:11

INTOXILYZER SERIAL # 80-002262 LOCATION DUI/HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Cheryl Straub

DPS Form Exh G-4 (Rev. 05-01).

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

287 cs

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002262
09/17/2009 10:17:29

154

Standard Lot#: 826001I
Last Changed By: C. MORAND #24219

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:18:34
Diagnostic Test	Pass	10:19:02
Air Blank	0.000	10:19:29
0.100 Cal Check	0.100	10:19:48
Air Blank	0.000	10:20:17
Subject Test	0.000	10:20:34
Air Blank	ABT*	10:20:41
Air Blank	0.000	10:21:09

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

387 CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002262
09/17/2009 10:22:09

154

Standard Lot#: 826001I
Last Changed By: C. MORAND #24219

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:23:33
Air Blank	IPS**	10:23:41

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

487 CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number : 80-002262
09/17/2009 10:24:20

Standard Lot#: 826001I
Last Changed By: C. MORAND #24219

154

QAS: C. STRAUB #50772
TUCSON PD

Operator: C. STRAUB #50772
TUCSON PD

Subject: SQAP,SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:25:12
Diagnostic Test	Pass	10:25:40
Air Blank	0.000	10:26:07
0.100 Cal Check	0.100	10:26:26
Air Blank	0.000	10:26:55
Subject Test	0.246*	10:30:03
Air Blank	0.000	10:30:36
Air Blank	0.000	10:31:03
Subject Test	XXX**	10:31:21
Air Blank	0.000	10:31:55

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department

527
CS

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002262
09/17/2009 10:32:26

Standard Lot#: 826001I
Last Changed By: C. MORAND #24219

QAS: C. STRAUB #50772
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:32:53
Cal Check	0.100	10:33:12
Air Blank	0.000	10:33:42

134

Operator's Signature

Pursuant to section 28-1337, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

677 CS

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

154 10-2-09

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department

DATE 09/17/09 TIME 10:11

Intoxilyzer Serial # 80-002262 Location DUI/HQ

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cheryl Straub

DPS Form Exh G-2 (Rev. 05-01)

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

717 cs

INTOXILYZER 8000
Location: ARIZONA
Serial Number: 80-002262
09/17/2009 10:11:42

Standard Lot#: 826001I
Last Changed By: C. MORAND #24219

QAS: C. STRAUB #50772
TUCSON PD

154

Test	g/210L	Time
Air Blank	0.000	10:12:10
Cal Check	0.101	10:12:28
Air Blank	0.000	10:12:58

Operator # Signature

Pursuant to section 28-1527, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department