



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

74

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002264 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 3.15.2007 Cal 1 0.101 AC Cal 2 0.101 AC

Between the dates of 2.15.2007 and 3.15.2007 the instrument was operating properly and accurately.


The instrument was taken out of service for repair on —.
The instrument was returned to service on —.

Terry S. Gallegos
Terry S. Gallegos

April 6, 2007

ADPS QAS Permit # 33169

SUBSCRIBED AND SWORN to before me this 6th day of April, 2007.



Kathy A. Kinney
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1057
Tph
ms

QA SPECIALIST Ferry Gallegos 28947 AGENCY Tucson Police Department
DATE 3.15.2007 TIME 1824
INTOXILYZER SERIAL # 80-002264 LOCATION Traffic

() 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 101 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Ferry Gallegos

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
03/15/2007

18:26:21

2 of 7
Dym

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:27:12
Diagnostic Test	Pass	18:27:40
Air Blank	0.000	18:28:07
0.100 Cal Check	0.102	18:28:26
Air Blank	0.000	18:28:55
Subject Test	0.000	18:29:20
Air Blank	0.000	18:29:47
Five Minute Wait	ABT*	
Air Blank	0.000	18:30:28

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
03/15/2007

18:30:33

3087
TGM

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:31:25
Diagnostic Test	Pass	18:31:52
Air Blank	RFI*	18:32:07
Air Blank	IPS**	18:32:17

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

4 of 7
B4.m

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
03/15/2007 18:32:35

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Operator: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Subject: SQAP, SQAP
DOB: 11/11/1111
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:33:22
Diagnostic Test	Pass	18:33:49
Air Blank	0.000	18:34:17
0.100 Cal Check	0.101	18:34:35
Air Blank	0.000	18:35:05
Subject Test	0.059*	18:38:13
Air Blank	0.000	18:38:45
Air Blank	0.000.	18:39:12
Subject Test	XXX**	18:39:50
Air Blank	0.000	18:40:24

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-002264
03/15/2007 18:40:57

50F 7
154m

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	18:41:25
Cal Check	0.101	18:41:43
Air Blank	0.000	18:42:13



Operator's Signature

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6 of 7
12/1/07

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Terry Gallegos 28887 AGENCY Tucson Police Department
DATE 3.15.2007 TIME 1824
Intoxilyzer Serial # 80-002264 Location Traffic

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-002264
03/15/2007 18:24:38

7077
P4MP

Standard Lot#: 629202I
Last Changed By: T. GALLEGOS #28847

QAS: T. GALLEGOS #28847
TUCSON PD CRIME LAB

Test	g/210L	Time
Air Blank	0.000	18:25:05
Cal Check	0.101	18:25:24
Air Blank	0.000	18:25:53



Operator's Signature