



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

TJ
3-17-08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002264 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 3-13-08 Cal 1 0.103 AC Cal 2 0.101 AC

Between the dates of 2-14-08 and 3-13-08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.


Clare Conley

Clare L. Conley

3-14-08

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 14, day of March, 2008.



Kathy A Kinney
Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

187
on
by
3.17.08

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 3-13-08 TIME 10:24

INTOXILYZER SERIAL # 80-002264 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 101 AC

Instrument is operating properly and accurately. YES NO

COMMENTS None.

SIGNATURE Clare Conley

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
03/13/2008

10:26:31

2867

cu

124

Standard Lot#: 7254021
Last Changed By: J. HAWKINS #42401

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:27:40
Diagnostic Test	Pass	10:28:08
Air Blank	0.000	10:28:35
0.100 Cal Check	0.103	10:28:54
Air Blank	0.000	10:29:23
Subject Test	0.000	10:29:42
Air Blank	ABT*	10:29:49
Air Blank	0.000	10:30:17

*Sequence Aborted

Not a Successfully Completed Test Sequence

Pursuant to section 16-1521, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
03/13/2008

10:30:30

347
cu
174

Standard Lot#: 7254021
Last Changed By: J. HAWKINS #42401

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:31:08
Air Blank	IPS**	10:31:16

*RFI Detect
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002264
03/13/2008 10:35:34

4577
cu
TSM

Standard Lot#: 7254021
Last Changed By: J. HAWKINS #42401

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST, SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:36:33
Diagnostic Test	Pass	10:37:00
Air Blank	0.000	10:37:27
0.100 Cal Check	0.102	10:37:46
Air Blank	0.000	10:38:16
Subject Test	0.123*	10:41:24
Air Blank	0.000	10:41:56
Air Blank	0.000	10:42:23
Subject Test	XXX**	10:42:43
Air Blank	0.000	10:43:17

*Deficient Sample
**Invalid Sample
Begin new deprivation period

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-002264

03/13/2008

10:43:51

527
CW
TS4

Standard Lot#: 7254021

Last Changed By: J. HAWKINS #42401

QAS: C. CONLEY #34517

TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:44:18
Cal Check	0.101	10:44:37
Air Blank	0.000	10:45:07

Claire Conley

Operator's Signature

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

6577
Cu
754

QA Specialist Clare Conley AGENCY Tucson Police Department
DATE 3-13-08 TIME 10:24
Intoxilyzer Serial # 80-202264 Location DUI HQ

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 103 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-002264

03/13/2008

10:24:24

7867
Cu
734

Standard Lot#: 7254021

Last Changed By: J. HAWKINS #42401

QAS: C. CONLEY #34517

TUCSON PD

Test	g/210L	Time
Air Blank	0.000	10:24:51
Cal Check	0.103	10:25:10
Air Blank	0.000	10:25:39

Clare Conley
Operator's Signature