



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory
1306 W. Miracle Mile, Tucson, AZ 85705

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002264 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TSK
5.17.2016

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4/28/2016 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 3/31/2016 and 4/28/2016 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on _____.
The instrument was returned to service on _____.

William L Loomis

William L. Loomis


5/3/2016

ADPS QAS Permit # 33720

SUBSCRIBED AND SWORN to before me this 3rd, day of May, 2016.

Elma Escarcega

Notary Public



*Includes documents specified in A.A.C. R13-10-104.A.5

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

1 of 7
WW
TSY 5.27.2016

QA SPECIALIST W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 4/28/2016 TIME 16:05
INTOXILYZER SERIAL # 80-002264 LOCATION CRIME LAB

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
- (✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
- (✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (✓) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 098 AC

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Willie Loomis

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002264
Core Version: 8105.48
04/28/2016

16:07:31

Standard Lot#: 25414100A4
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:08:49
Diagnostic Test	Pass	16:09:24
Air Blank	0.000	16:09:51
0.100 Cal Check	0.098	16:10:12
Air Blank	0.000	16:10:42
Subject Test	0.000	16:11:10
Air Blank	0.000	16:11:37
Wait	ABT*	
Air Blank	0.000	16:12:11

*Sequence Aborted

2 of 7
WV
Tdy

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002264
Core Version: 8105.48
04/28/2016

16:12:30

Standard Lot#: 25414100A4
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:13:19
Diagnostic Test	Pass	16:13:54
Air Blank	0.000	16:14:21
0.100 Cal Check	RFI*	16:14:35
Air Blank	IPS**	16:14:46

*RFI Detect
**Improper Sample

3 of 7
WLL
154

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002264
Core Version: 8105.48
04/28/2016

16:14:57

Standard Lot#: 25414100A4
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Subject: SQAP, SQAP
DOB: 11/11/1987
Sex: M Weight: 222

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:16:14
Diagnostic Test	Pass	16:16:49
Air Blank	0.000	16:17:17
0.100 Cal Check	0.099	16:17:38
Air Blank	0.000	16:18:07
Subject Test	0.173*	16:21:15
Air Blank	0.000	16:21:48
Air Blank	0.000	16:22:16
Subject Test	XXX**	16:22:40
Air Blank	0.000	16:23:15

*Deficient Sample

**Invalid Sample

Begin new deprivation period

4 of 7
WLL
T34

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002264
Core Version: 8105.48
04/28/2016 16:23:31

Standard Lot#: 25414100A4
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	16:24:00
0.100 Cal Check	0.098	16:24:21
Air Blank	0.000	16:24:51

Preventative Maintenance
Performed -- Timer Reset

5 of 7
WLL
TS4

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

6 of 7
WLL
TSG 5.17.2016

QA Specialist W. LOOMIS 53545 AGENCY Tucson Police Department
DATE 4/28/2016 TIME 16:05
Intoxilyzer Serial # 80-002264 Location CRIME LAB

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Willie Loomis

DPS Form Exh G-2 (Rev. 05-01)

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002264
Core Version: 8105.48
04/28/2016 16:05:55

Standard Lot#: 25414100A4
Last Changed By: T. GALLEGOS #28847

QAS: W. LOOMIS #53545
TUCSON PD

Operator: W. LOOMIS #53545
TUCSON PD

Test	g/210L	Time
Air Blank	0.000	16:06:24
0.100 Cal Check	0.099	16:06:45
Air Blank	0.000	16:07:15

Preventative Maintenance
Performed -- Timer Reset

7 of 7
WLL
TS4