CERTIFICATION by CUSTODIAN of RECORDS

*Tucson Police Department Crime Laboratory*
*270 S. Stone Avenue Tucson, AZ 85701-1917*

SC
11/21/08

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002265 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 11-13-08 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 10-17-08 and 11-13-08 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.
The instrument was returned to service on N/A.

Clare Conley

Clare L. Conley

11-20-08

ADPS QAS Permit # 33314

SUBSCRIBED AND SWORN to before me this 20th, day of November, 2008.

A.M. Hernandez

Notary Public

OFFICIAL SEAL
A.M. HERNÁNDEZ
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Mar. 17, 2012

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

1 of 7
Cu
SC
11/21/08

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Clare Conley AGENCY Tucson Police Department

DATE 11-13-08 TIME 0756

INTOXILYZER SERIAL # 80-002265 LOCATION ODD

( ✓ ) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

( ✓ ) 1. Clock time check.
( ✓ ) 2. Date check.

OPERATIONAL TESTS

( ✓ ) 1. Alcohol-free subject test result 0. 000 AC.
( ✓ ) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
( ✓ ) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
( ✓ ) 4. Standard Calibration Check standard 0. 100 AC
Result: 0. 100 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS None.

SIGNATURE Clare Conley

DPS Form Exh G-4 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

cy 2
Cu ~~H~~ 5 of 7
Cu

SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002265
11/13/2008 07:58:40

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 07:59:30 |
| Diagnostic Test | Pass | 07:59:57 |
| Air Blank | 0.000 | 08:00:25 |
| 0.100 Cal Check | 0.101 | 08:00:43 |
| Air Blank | 0.000 | 08:01:13 |
| Subject Test | 0.000 | 08:01:33 |
| Air Blank | ABT* | 08:01:40 |
| Air Blank | 0.000 | 08:02:08 |

*Sequence Aborted

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

cu cu cu
3 ~~4~~ 2 of 7
cu

SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002265
11/13/2008 08:02:13

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | RFI* | 08:02:46 |
| Air Blank | IPS** | 08:02:54 |

*RFI Detect
**Improper Sample

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

cn 4
cn 3 of 7
cn

SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number : 80-002265
11/13/2008 08:03:14

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

QAS: C. CONLEY #34517
TUCSON PD

Operator: C. CONLEY #34517
TUCSON PD

Subject: TEST,SQAP
DOB: 11/11/1111
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:04:03 |
| Diagnostic Test | Pass | 08:04:31 |
| Air Blank | 0.000 | 08:04:58 |
| 0.100 Cal Check | 0.101 | 08:05:17 |
| Air Blank | 0.000 | 08:05:46 |
| Subject Test | 0.293* | 08:08:55 |
| Air Blank | 0.000 | 08:09:27 |
| Air Blank | 0.000 | 08:09:54 |
| Subject Test | XXX** | 08:10:21 |
| Air Blank | 0.000 | 08:10:55 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7
Cu

SC

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-002265
11/13/2008 08:11:11

---

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

---

QAS: C. CONLEY #34517
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:11:38 |
| Cal Check | 0.100 | 08:11:57 |
| Air Blank | 0.000 | 08:12:27 |

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

b6b7
CW
SC

QA Specialist Clare Conley AGENCY Tucson Police Department

DATE 11-13-08 TIME 0756

Intoxilyzer Serial # 80-002265 Location ODD

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Clare Conley

DPS Form Exh G-2 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

INTOXILYZER 8000
Location: TUCSON POLICE DEPT
Serial Number: 80-002265
11/13/2008 07:56:08

---

Standard Lot#: 735401G
Last Changed By: S. CIONE #50616

---

QAS: C. CONLEY #34517
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 07:56:35 |
| Cal Check | 0.101 | 07:56:54 |
| Air Blank | 0.000 | 07:57:23 |

7 of 7
CW
SC

Clare Conley
Operator's Signature

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.