# CERTIFICATION by CUSTODIAN of RECORDS

*Tucson Police Department Crime Laboratory*
*1306 W. Miracle Mile, Tucson, AZ 85705*

TSG
10-31-2013

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002265 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 10/10/13 Cal 1 0.099 AC Cal 2 0.098 AC

Between the dates of 9/12/13 and 10/10/13 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on -.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione

10/21/13

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 21st, day of October, 2013.

Edna Escarcega

Notary Public

OFFICIAL SEAL
ELMA ESCARCEGA
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Nov. 3, 2015

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

1 of 7 sc
TSH
10-31-2013

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sarah Cione AGENCY Tucson Police Department

DATE 10/10/13 TIME 1054

INTOXILYZER SERIAL # 80-002265 LOCATION DUI HQ

(✓) 1. Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC
Result: 0.098 AC

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS _____

SIGNATURE Sarah Cione

DPS Form Exh G-4 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

2 of 7 SC
BH

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002265
Core Version: 8105.48
10/10/2013 11:03:20

---

Standard Lot#: 30412100A1
Last Changed By: S. CIONE #50616

---

QAS: S. CIONE #50616
TUCSON PD

---

Operator: S. CIONE #50616
TUCSON PD

---

Subject: SQAP,SQAP
DOB: 11/11/1980
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:04:04 |
| Diagnostic Test | Pass | 11:04:39 |
| Air Blank | 0.000 | 11:05:06 |
| 0.100 Cal Check | 0.099 | 11:05:27 |
| Air Blank | 0.000 | 11:05:57 |
| Subject Test | 0.000 | 11:06:14 |
| Air Blank | 0.000 | 11:06:42 |
| Wait | ABT* | |
| Air Blank | 0.000 | 11:07:49 |

*Sequence Aborted

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

3 of 7
SC
TSH

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002265
Core Version: 8105.48
10/10/2013 11:07:54

---

Standard Lot#: 30412100A1
Last Changed By: S. CIONE #50616

---

QAS: S. CIONE #50616
TUCSON PD

---

Operator: S. CIONE #50616
TUCSON PD

---

Subject: SQAP,SQAP
DOB: 11/11/1980
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:08:49 |
| Diagnostic Test | Pass | 11:09:24 |
| Air Blank | 0.000 | 11:09:52 |
| 0.100 Cal Check | 0.099 | 11:10:13 |
| Air Blank | 0.000 | 11:10:43 |
| Subject Test | RFI* | 11:11:36 |
| Air Blank | IPS** | 11:11:43 |

*RFI Detect
**Improper Sample

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

4 of 7 SC
TSH

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002265
Core Version: 8105.48
10/10/2013 11:17:04

---

Standard Lot#: 30412100A1
Last Changed By: S. CIONE #50616

---

QAS: S. CIONE #50616
TUCSON PD

---

Operator: S. CIONE #50616
TUCSON PD

---

Subject: SQAP,SQAP
DOB: 11/11/1980
Sex: F Weight: 111

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:18:03 |
| Diagnostic Test | Pass | 11:18:38 |
| Air Blank | 0.000 | 11:19:05 |
| 0.100 Cal Check | 0.098 | 11:19:26 |
| Air Blank | 0.000 | 11:19:56 |
| Subject Test | 0.158* | 11:23:04 |
| Air Blank | 0.000 | 11:23:35 |
| Air Blank | 0.000 | 11:24:03 |
| Subject Test | XXX** | 11:24:31 |
| Air Blank | 0.000 | 11:25:04 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

5 of 7 SC
154

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002265
Core Version: 8105.48
10/10/2013 11:25:40

---

Standard Lot#: 30412100A1
Last Changed By: S. CIONE #50616

---

QAS: S. CIONE #50616
TUCSON PD

---

Operator: S. CIONE #50616
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:26:08 |
| 0.100 Cal Check | 0.098 | 11:26:29 |
| Air Blank | 0.000 | 11:26:59 |

Preventative Maintenance
Performed -- Timer Reset

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY
A.A.C. R13-10-104(A)

6 of 7
SC
TSH
10.31.2013

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Sarah Cione AGENCY Tucson Police Department

DATE 10/10/13 TIME 1054

Intoxilyzer Serial # 80-002265 Location DUI HQ

( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

( ✓ ) 4. Air blank completed.

( ✓ ) 5. Calibration check completed. Test results 0.099 AC.

( ✓ ) 6. Air blank completed.

( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Sarah Cione

DPS Form Exh G-2 (Rev. 05-01)

*Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.*

INTOXILYZER 8000
Location: TUCSON PD
Serial Number: 80-002265
Core Version: 8105.48
10/10/2013 10:54:26

---

Standard Lot#: 30412100A1
Last Changed By: S. CIONE #50616

---

QAS: S. CIONE #50616
TUCSON PD

---

Operator: S. CIONE #50616
TUCSON PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:54:55 |
| 0.100 Cal Check | 0.099 | 10:55:15 |
| Air Blank | 0.000 | 10:55:45 |

Preventative Maintenance
Performed -- Timer Reset

7 of 7 SC
TSH

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.