CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 8000 Serial Number 80-002345 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TSA 4.11.2008

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 4/7/08 Cal 1 N/A AC Cal 2 N/A AC

Between the dates of 3/13/08 and N/A the instrument was operating properly and accurately.

The instrument was taken out of service for repair on 4/7/08.
The instrument was returned to service on -.

Sarah Cione

Sarah L. Cione

4/11/08

ADPS QAS Permit # 33355

SUBSCRIBED AND SWORN to before me this 11, day of April, 2008.

OFFICIAL SEAL
KATHY A. KINNEY
NOTARY PUBLIC-ARIZONA
PIMA COUNTY
My Comm. Exp. Sept. 4, 2008

Kathy A Kinney

Notary Public

*Includes documents specified in A.A.C. R13-10-104.A.5

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

TUCSON POLICE DEPARTMENT
CRIME LABORATORY

1 of 2 sc

INTOXILYZER SERVICE RECORD

TJ4
4.11-2008

Date: 4/7/08

QAS: Sarah Cione

Intoxilyzer Serial Number: 80-002345

Reason For Service: DSP Failure during Diagnostic Check
Officer Morgan brought intox to lab after Diagnostic Failure

Date Out of Service: 4/7/08

Calibration Check completed prior to removal from service

Date: –

SQAP completed prior to removal from service

Date: –

Corrective Action:

☒ Sent to CMI, Inc.

CoBRA data downloaded

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

CMI INC.

Return Material Authorization Form

2 of 2 SC

This form MUST be completed and enclosed with item(s) being returned. PY
Failure to compete and return this form may cause delays in crediting / repairing.

1. **Item(s) Returned:** ☑ Instrument ☐ Supplies ☐ Other
(Note: please ship items in their original shipping container(s) or a similar protective box.)

2. **Instrument Model:** Intox 8000 **Serial Number** 80-002345

3. **Bill To Address:**
Finance / Accounts Payable
P.O. Box 27450
Tucson, AZ 85726-7450

Ship To Address: (NO PO BOXES)
Tucson Police Department
Crime Lab
270 S. Stone Ave
Tucson, AZ 85701-1917

Customer # 105075
Name Terry Gallegos
Phone: (520) 791-4499 x1443 **Fax:** (520) 791-4642
Email: Terry.Gallegos@TucsonAz.gov

4. **Reason for Return:** *Note - If you are sending an item for repair, please give a detailed description of the problem. Please list any special instructions that you may have concerning this return. Note - Returns for credit may be subject to a 15% restocking charge.*
DSP failure during initial Diagnostic Check

5. **Purchase Order Number** (attach a copy of P.O. if applicable) 20669

6. **I Authorize All Repairs:** ☒ Yes ☐ No

7. **I Authorize Repairs Up To:** $_____

8. **I Require An Estimate Regardless of Cost** ☐ Yes ☐ No
Note: An evaluation fee will apply to estimates that are not repaired.

9. **Authorized By:**
Terry Gallegos
Name (Please Print)
Terry Gallegos
Signature
4.10.2008
Date

Pursuant to Section 28.1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.